

# Defibtech DDU-2200 Fully-Automatic External Defibrillator



## User Manual

For comprehensive training on set-up,  
use and maintenance; source for  
complete technical specifications


**PRESS "ON"  
BUTTON**

**1**



**APPLY PADS**

**2**



**FOLLOW AED  
INSTRUCTIONS**

**3**



## Notices

Defibtech, L.L.C. shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Information in this document is subject to change without notice. Names and data used in any examples are fictitious unless otherwise noted.

## Limited Warranty

The "Limited Warranty" shipped with Defibtech AED products serves as the sole and exclusive warranty provided by Defibtech, L.L.C., with respect to the products contained herein.

## Copyright

Copyright © 2017 Defibtech, L.L.C.

All rights reserved. Copyright questions should be directed to Defibtech.

For contact information, refer to Chapter 12 of this manual.




## Tracking

U.S.A. federal regulations require Defibtech to maintain records for each AED it distributes (reference 21 CFR 821, Medical Device Tracking). These requirements also apply anytime there is a change in the AED's location, including if you move, sell, donate, give away, export or even throw it away. We depend on AED owners/users to contact us when these things happen to ensure the tracking information remains accurate in the event we need to share important product notices. If your location is outside the U.S.A., we ask you share your information for exactly the same reasons. To keep your information up to date, please visit [www.defibtech.com/register](http://www.defibtech.com/register).



**Federal Law (USA) restricts this device  
to sale by or on the order of a physician.**

# Contents

|          |  |           |
|----------|--|-----------|
| <b>1</b> | <b>Introduction to the DDU-2200 AED .....</b>  | <b>6</b>  |
| 1.1      | Overview .....   | 6         |
| 1.2      | The Defibtech DDU-2200 AED .....   | 7         |
| 1.3      | Usage .....  | 9         |
| <b>2</b> | <b>Dangers, Warnings, and Cautions .....</b>   | <b>10</b> |
| 2.1      |  Dangers .....  | 10        |
| 2.2      |  Warnings.....  | 10        |
| 2.3      |  Cautions ..... | 11        |
| <b>3</b> | <b>Setting Up the DDU-2200 AED.....</b>  | <b>13</b> |
| 3.1      | Overview .....   | 13        |
| 3.2      | Connecting the Defibrillation Pads .....   | 14        |
| 3.3      | Installing and Removing the Battery Pack .....   | 14        |
| 3.4      | Checking the DDU-2200 AED Status .....   | 15        |
| 3.5      | Installing the Defibtech Data Card (DDC Card) (optional) .....                                   | 16        |
| 3.6      | Completing the Installation .....  | 16        |
| 3.7      | Storing the DDU-2200 AED.....  | 16        |
| <b>4</b> | <b>Using the DDU-2200 AED.....</b>   | <b>17</b> |
| 4.1      | Overview .....   | 17        |
| 4.2      | Preparation .....  | 18        |
| 4.3      | Heart Rhythm Analysis .....  | 21        |
| 4.4      | Delivering the Shock.....  | 21        |
| 4.5      | CPR Period .....   | 21        |
| 4.6      | Post-Use Procedures.....   | 22        |
| 4.7      | Voice and Text Prompts .....   | 22        |
| 4.8      | Operational Environment.....   | 26        |
| <b>5</b> | <b>Maintenance and Troubleshooting.....</b>  | <b>27</b> |
| 5.1      | Routine Unit Maintenance.....  | 27        |
| 5.2      | Self-Tests.....  | 31        |
| 5.3      | Cleaning .....   | 31        |
| 5.4      | Storage.....   | 31        |
| 5.5      | Operator's Checklist .....   | 32        |
| 5.6      | Troubleshooting .....  | 33        |
| 5.7      | Repair .....   | 34        |

|           |  |           |
|-----------|--|-----------|
| <b>6</b>  | <b>Maintenance Mode .....</b>                | <b>35</b> |
| 6.1       | Overview .....                               | 35        |
| 6.2       | Navigation (in Maintenance Mode).....        | 35        |
| 6.3       | Entering Maintenance Mode.....               | 36        |
| 6.4       | AED Main Menu Screen .....                   | 36        |
| 6.5       | AED Status Screen.....                       | 36        |
| 6.6       | AED Maintenance Screen .....                 | 37        |
| 6.7       | AED Options Screen .....                     | 39        |
| 6.8       | Rescue Options Screen.....                   | 42        |
| 6.9       | Help Topics Screen .....                     | 43        |
| <b>7</b>  | <b>DDU-2200 AED Accessories.....</b>         | <b>44</b> |
| 7.1       | Defibrillation Pads.....                     | 44        |
| 7.2       | Battery Packs.....                           | 44        |
| 7.3       | Data Cards.....                              | 44        |
| 7.4       | USB Cable .....                              | 44        |
| <b>8</b>  | <b>Event Viewing .....</b>                   | <b>45</b> |
| 8.1       | DefibView.....                               | 45        |
| 8.2       | Defibtech Data Cards (DDC Cards) .....       | 45        |
| 8.3       | Downloading the Internal Data Log.....       | 45        |
| <b>9</b>  | <b>Technical Specifications.....</b>         | <b>46</b> |
| 9.1       | Defibtech DDU-2200 AED .....                 | 46        |
| 9.2       | Battery Packs.....                           | 50        |
| 9.3       | Self-Adhesive Defibrillation Pads.....       | 50        |
| 9.4       | Event Documentation .....                    | 50        |
| 9.5       | Defibtech Event Viewer .....                 | 50        |
| 9.6       | Recycling Information.....                   | 51        |
| 9.7       | Notice to European Union Customers .....     | 51        |
| <b>10</b> | <b>Electromagnetic Conformity.....</b>       | <b>52</b> |
| 10.1      | Guidance and Manufacturer's Declaration..... | 52        |
| <b>11</b> | <b>Glossary of Symbols.....</b>              | <b>55</b> |
| <b>12</b> | <b>Contacts .....</b>                        | <b>57</b> |

# 1 Introduction to the DDU-2200 AED

This User Manual provides information to guide operators in the use and maintenance of the Defibtech DDU-2200 Automated External Defibrillator (AED) and its accessories. It includes comprehensive training on set-up, use, and maintenance and is the source for complete technical specifications.

This chapter includes intended use, an overview of the AED, a discussion of when it should and should not be used, and information on operator training.

## 1.1 Overview

The DDU-2200 AED is a Fully-Automatic External Defibrillator that will shock a patient automatically without additional interaction by the operator if a shock is required. It is designed to be easy to use, portable, and battery powered. Push-button controls and indicators include an ON/OFF button, three softkey buttons, and a SHOCK Required Indicator LED. Voice prompts, text prompts, and a display screen with visual prompts provide a simple interface for the operator. The DDU-2200 AED is capable of recording event information, including ECG, audio data (optional), and SHOCK/NO-SHOCK recommendations.

The DDU-2200 AED performs the following tasks:

- Prompts the operator, through audio, text, and video prompts, to prepare the patient for treatment.
- Automatically analyzes the patient's ECG.
- Determines whether a shockable rhythm is present.
- Charges the capacitor.
- Automatically delivers a shock, without any user intervention, once the device has determined a shock is advised.
- Provides instructions to perform CPR.
- Repeats the process, if needed.

The DDU-2200 AED uses two self-adhesive, single-use, non-sterile defibrillation pads (also known as electrode pads or pads) to monitor ECG signals and, if advised, to deliver defibrillation energy to the patient. These pads are provided in a single-use, disposable package that can be preconnected to the AED. The pads package is labeled with an expiration date. The DDU-2200 AED determines proper pad-to-patient contact by measuring the impedance between the two pads.

The DDU-2200 AED user interface is clear and concise. It has push-button controls and a display screen. Easily understandable voice messages and text and video prompts guide the operator through the use of the unit. The device communicates the status of the AED and of the patient to the operator.

Defibrillation energy is delivered as an impedance compensated biphasic truncated exponential waveform. The device delivers 150 Joules of defibrillation energy (into a 50-ohm load) when using adult defibrillation pads and 50 J of defibrillation energy (into a 50-ohm load) when using child/infant pads (also known as pediatric defibrillation pads). Energy delivered does not change significantly with patient impedance, although the duration of the generated waveform will vary.

Defibrillation and AED operating power is supplied with a replaceable (non-rechargeable) battery pack that provides for long standby life and low maintenance operation. Each battery pack is marked with an expiration date.


The DDU-2200 AED records event documentation internally and, optionally, on a Defibtech Data Card (DDC card). The optional DDC card inserts into a slot in the AED and enables the AED to record event documentation and, optionally, audio data onto the card. Audio recording is selectable through configuration settings. Event documentation stored internally can be downloaded onto a DDC card for review.

A USB port is provided to perform maintenance and data recovery. The USB interface allows connection to a personal computer. Defibtech PC maintenance software helps support event downloading and unit maintenance operations.


## 1.2 The Defibtech DDU-2200 AED

---

- A. Speaker.** The speaker projects the voice prompts when the DDU-2200 AED is on. The speaker also emits a “beep” when the unit is off and has detected a condition that requires attention from the user or needs servicing.
- B. SHOCK Required Indicator.** Flashes when a shock is recommended and the unit has charged and is to deliver a shock. **Do not touch the patient while this indicator is flashing.**
- C. Display Screen.** Color display panel used to display text and video prompts, messages, indicators for rescue, unit status, and maintenance operations. The display screen provides visual prompts, including CPR coaching.
- D. ON/OFF Button.** This button is used to turn the DDU-2200 AED on and off.
- E. Pads Connector Socket.** The pads connector (item N) is inserted into this socket.
- F. Active Status Indicator (ASI).** The ASI indicates the current status of the AED. This indicator flashes green to indicate the unit has passed its last self-test and is ready for use. It flashes red to indicate unit needs attention from the user or needs servicing.
- G. Softkey Buttons.** Three context sensitive softkey buttons are used to navigate menus or select actions.
- H. USB Port.** The USB port is provided to perform data recovery and maintenance. Not to be used during rescue operation.
- I. Defibtech Data Card (DDC card).** This optional plug-in card provides enhanced storage capabilities to the AED.
- J. USB and Defibtech Data Card (DDC card) Access Door.** Behind the access door is the USB connector port and Defibtech Data Card (DDC card) slot.
- K. Battery Pack Opening.** This opening is where the battery pack is inserted into the unit.
- L. Battery Pack Eject Release Latch.** This release latch releases the battery pack from the DDU-2200 AED.
- M. Pad Storage Area.** The pad storage area is found on the back of the AED allowing the pads to be stored in a pre-connected state for rapid deployment during an emergency.
- N. Pads Connector.** This connector attaches the patient pads to the unit at the pads connector socket (item E).
- O. Defibrillation Pads.** The defibrillation pads are pads that are placed on the patient. The pads should be stored in the pad storage area (item M) on the back of the unit.
- P. Defibrillation Pads Expiration Date (back side).** The defibrillation pads expiration date is located on the back side of the pads package. Do not use the pads after the printed date has passed.
- Q. Battery Pack.** The battery pack provides a replaceable main power source for the DDU-2200 AED.



FRONT OF AED




BACK OF AED

N. Pads Connector




O. Defibrillation Pads Expiration Date (on back of package)



P. Defibrillation Pads

Q. Battery Pack



BATTERY PACK



## 1.3 Usage

---

### Indications

The DDU-2200 Automated External Defibrillator (AED) is indicated for use on victims of sudden cardiac arrest (SCA) who are:

- Unconscious and unresponsive
- Not breathing or not breathing normally

For patients under 8 years old, or weighing less than 55 lbs (25 kg), use child/infant defibrillation pads, if available. Do not delay therapy to determine exact age or weight. Apply the pads as shown for a child/infant and use the AED.

Federal Law (USA) restricts this device to sale by or on the order of a physician.

### Contraindications

None.

### Operator Training Requirements

In order to safely and effectively operate the DDU-2200 AED, a person shall have met the following requirements:

- Defibtech DDU-2200 AED and/or defibrillation training as required by local, state, provincial, or national regulations.
- Any additional training as required by the authorizing physician.
- Thorough knowledge and understanding of the material presented in this User Manual.

## 2 Dangers, Warnings, and Cautions

This chapter includes a list of danger, warning, and caution messages that relate to the Defibtech DDU-2200 AED and its accessories. Many of these messages are repeated elsewhere in this User Manual and on the DDU-2200 AED or accessories. The entire list is presented here for convenience.

### 2.1 **DANGERS:**

**Immediate hazards that will result in serious personal injury or death.**

---

- Hazardous electrical output. This equipment is for use only by qualified personnel.
- Possible fire or explosion. Do not use in the presence of flammable gases or anesthetics. Use care when operating this device close to oxygen sources (such as bag-valve-mask devices or ventilator tubing). Turn off gas source or move source away from patient during defibrillation, if necessary.
- The DDU-2200 AED has not been evaluated or approved for use in hazardous locations as defined in the National Electric Code standard. In compliance with IEC classification, the DDU-2200 AED is not to be used in the presence of flammable substance/air mixtures.

### 2.2 **WARNINGS:**

**Conditions, hazards, or unsafe practices that may result in serious personal injury or death.**

---

- Not intended to be used in an environment with high-frequency electrosurgical equipment.
- Improper use can cause injury. Use the DDU-2200 AED only as instructed in the User Manual. The DDU-2200 AED delivers electrical energy that can potentially cause death or injury if it is used or discharged improperly.
- Improper maintenance can cause the DDU-2200 AED not to function. Maintain the DDU-2200 AED only as described in the User Manual and Operating Guide. The AED contains no user-serviceable parts — do not take the unit apart.
- No modification of this equipment is allowed.
- Electrical Shock Hazard. Dangerous high voltages and currents are present. Do not open unit, remove cover (or back), or attempt repair. There are no user serviceable components in the DDU-2200 AED. Refer servicing to qualified service personnel.
- DBP-2003 and DBP-2013 battery packs are not rechargeable. Any attempt to recharge these battery packs may result in fire or explosion.
- Do not immerse battery pack in water or other liquids. Immersion in fluids may result in fire or explosion.
- Do not let fluids get into the DDU-2200 AED. Avoid spilling fluids on the AED or its accessories. Spilling fluids into the DDU-2200 AED may damage it or cause a fire or shock hazard.
- Do not sterilize the DDU-2200 AED or its accessories.
- Use only Defibtech disposable self-adhesive defibrillation pads, battery packs, and other accessories supplied by Defibtech or its authorized distributors. Substitution of non-Defibtech approved accessories may cause the device to perform improperly.
- Do not open sealed pads package until pads are to be used.
- Do not touch the patient during defibrillation. Defibrillation current can cause operator or bystander injury.
- Do not allow pads to touch metal objects or equipment in contact with the patient. Do not touch equipment connected to the patient during defibrillation. Disconnect other electrical equipment from the patient before defibrillation.
- Do not shock with defibrillation pads touching each other. Do not shock with gel surface exposed.
- Do not allow defibrillation pads to touch each other, or to touch other ECG electrodes, lead wires, dressings, transdermal patches, etc. Such contact can cause electrical arcing and patient skin burns during defibrillation and may divert defibrillating energy away from the heart.
- The defibrillation pads are intended for one-time use only and must be discarded after use. Reuse can lead to potential cross infection, improper performance of the device, inadequate delivery of therapy, and/or injury to the patient or operator.

*WARNINGS (continued)*

- Avoid contact between parts of the patient's body and conductive fluids such as water, gel, blood or saline, and metal objects, which may provide unwanted pathways for defibrillating current.
- Disconnect all non-defibrillator proof equipment from the patient before defibrillation to prevent electrical shock hazard and potential damage to that equipment.
- Aggressive or prolonged CPR to a patient with defibrillation pads attached can cause damage to the pads. Replace the defibrillation pads if they become damaged during use.
- Possible Radio Frequency (RF) interference from RF devices such as cellular phones and two-way radios can cause improper AED operation. Normally using a cell phone near the AED should not cause a problem; however, a distance of 2 meters (6 feet) between RF devices and the DDU-2200 AED is recommended.
- CPR during analysis can cause incorrect or delayed diagnosis by the patient analysis system.
- Handling or transporting the patient during ECG analysis can cause incorrect or delayed diagnosis, especially if very low amplitude or low frequency rhythms are present. If the patient is being transported, stop vehicle before beginning ECG analysis.
- In patients with cardiac pacemakers, the DDU-2200 AED may have reduced sensitivity and not detect all shockable rhythms. If you know the patient has an implanted pacemaker, do not place electrodes directly over an implanted device.
- During defibrillation, air pockets between the skin and defibrillation pads can cause patient skin burns. To help prevent air pockets, make sure self-adhesive defibrillation pads completely adhere to the skin. Do not use dried out or expired defibrillation pads.
- Defibrillation may cause skin burns around the defibrillation pads area.
- User-initiated and automatic self-tests are designed to assess the DDU-2200 AED's readiness for use. However, no degree of testing can assure performance or detect abuse, damage, or a defect that occurred after the most recent test is completed.
- Use of damaged equipment or accessories may cause the device to perform improperly and/or result in injury to the patient or operator.
- Follow voice prompts if the LCD screen becomes blank or unreadable.
- It may be possible for the AED to not detect a shockable rhythm, not deliver a shock to a shockable rhythm or not deliver the intended energy during defibrillation.
- It may be possible that the AED recommends a shock for a non-shockable rhythm, and if a shock is delivered, VF or cardiac arrest may occur.
- Even if defibrillation occurs, the sudden cardiac arrest event may not result in survival.
- Defibrillation may cause myocardial damage or post-shock dysfunction.

### **2.3 CAUTIONS:**

**Conditions, hazards, or unsafe practices that may result in minor personal injury, damage to the DDU-2200 AED, or loss of data.**

---

- Follow all battery pack labeling instructions. Do not install battery packs after the expiration date.
- Follow all defibrillation pad label instructions. Use defibrillation pads prior to their expiration date. Do not re-use defibrillation pads. Discard defibrillation pads after use (in the event of suspected pad malfunction, return pads to Defibtech for testing).
- The defibrillation pads should not be in continuous contact with the patient's skin for more than 24 hours.
- Allergic dermatitis or a minor skin rash may result in patients that are sensitive to the materials used for the defibrillation pads. Remove the defibrillation pads from the patient as soon as practical.
- Recycle or dispose of lithium battery packs in accordance with local, state, provincial, and/or national regulations. To avoid fire and explosion hazard, do not burn or incinerate the battery pack. Do not crush.

*CAUTIONS (continued)*


- Use and store the DDU-2200 AED only within the range of environmental conditions specified in the technical specifications.
- If possible, disconnect the DDU-2200 AED from the patient prior to use of other defibrillators.
- Do not connect the DDU-2200 AED to a PC or other device (using the USB port) while the unit's electrodes are still connected to the patient.
- Using non-Defibtech Data Cards (DDC cards) may damage the unit and will void the warranty.
- Although the DDU-2200 AED is designed for a wide variety of field use conditions, rough handling beyond specifications may result in damage to the unit.
- Federal Law (USA) restricts this device to sale by or on the order of a physician.

# 3 Setting Up the DDU-2200 AED

This chapter describes the steps required to make your Defibtech DDU-2200 AED operational. The DDU-2200 AED is designed to be stored in a "ready" state. This chapter tells you how to make the device ready, so that if and when you need it, few steps are required to begin using the device.

## 3.1 Overview

The following components and accessories are included with your DDU-2200 AED. Replacement and other accessories are detailed in the "DDU-2200 AED Accessories" chapter. Before getting started, identify each component and ensure that your package is complete.



### 3.2 Connecting the Defibrillation Pads

The DDU-2200 AED defibrillation pads are supplied sealed in a package with the connector and part of the cable exposed.



**DO NOT** open the sealed pads package until the pads are to be used. The packaging should be opened only immediately prior to use, otherwise the pads may dry out and become non-functional.

**Note:** The DDU-2200 AED is designed to be stored with the pads connector already installed. This simplifies the procedure for deploying and operating the device in an emergency.

First, check to ensure that the pads package has not expired. The expiration date is printed on the pad pouch and is also reported on the AED status screen. Do not use pads past expiration date. Discard expired pads.

|  |  |
|--|--|
|  | <p>Insert the connector end of the defibrillation pad cable into the pads connector socket on the top-left corner of the DDU-2200 AED as shown. Insert pads connector firmly until it is fully seated in the unit. The connector will only fit in one way – if the connector does not fit, rotate the connector before trying again.</p> <p>The connected pads package should then be stored in the pad storage slot in the back of the DDU-2200 AED. After connecting the pads connector to the unit, push the pads package, rounded end first, with the pictures on the package facing out, into the pad holder compartment on the back of the AED. When the pads package is fully inserted, press the pad cable into the groove in the back of the unit to hold the cable in place and tuck any excess cable behind the pads package.</p> |
|--|--|



The pads are intended for one time use only and must be discarded after use or if the package has been opened or damaged.

### 3.3 Installing and Removing the Battery Pack

The battery pack provides power to the DDU-2200 AED. Do not install the battery pack after the expiration date printed on the label. The supplied battery pack is non-rechargeable.

|  |   |
|--|---|
|  | <p>Before inserting the battery pack into the DDU-2200 AED, ensure that the battery pack opening in the back of the AED is clean and clear of any foreign objects. Insert the battery pack into the opening in the back of the AED. Push the pack all the way in until the latch clicks. The battery pack will only fit in one way – if the battery pack does not fit, rotate the battery pack before trying again. Once fully inserted, the battery pack surface should be flush with the back of the AED.</p> <p>Within moments of insertion the DDU-2200 AED will turn on and run a battery pack insertion self-test.* When the test is completed, the unit will report the status of the battery pack and shut down. Afterwards, the Active Status Indicator (ASI), adjacent to the ON/OFF button of the DDU-2200 AED, will periodically flash. If the indicator flashes green, the AED and battery pack are ready for use. If the indicator flashes red, is solid red, or there is no flashing light, the AED requires service (refer to Section 3.4, "Checking the DDU-2200 AED Status," for more details on the meaning of the indicator).</p> |
|--|---|

**\*Note:** The battery pack must have been removed from the unit for at least 10 seconds for the battery pack self-test to be performed automatically.


To remove the battery pack, push the battery pack eject release latch. After the battery pack is partially ejected, pull the battery pack out.


### 3.4 Checking the DDU-2200 AED Status

#### Active Status Indicator (ASI)

Once a fully-functional battery pack is installed in the DDU-2200 AED, an LED indicator located to the right of the ON/OFF button actively indicates unit status. If the unit is ready for use, the Active Status Indicator (ASI) will flash green. Ready for use means that the DDU-2200 AED has passed the most recent self-test (scheduled or user initiated). If the unit needs service, the ASI will flash red. When the ASI flashes red, the unit will also “beep” periodically to call attention to itself. The ASI also uses a distinct flash pattern to assist people with color blindness: green will flash a single flash and red will flash a double flash.

The ASI is powered by the battery pack. If the battery pack has been completely discharged or is not installed in the unit, the active status indication will be off. In this case, immediately replace the battery pack or reinsert it into the unit to restore active status indication.

|   |  |
|---|--|
|  <p><b>Active Status Indicator (ASI)</b></p> | <ul style="list-style-type: none"> <li>• <b>Flashing Green:</b> The DDU-2200 AED is OFF and ready for use.</li> <li>• <b>Solid Green:</b> The DDU-2200 AED is ON and ready for use.</li> <li>• <b>Flashing or Solid Red:</b> The DDU-2200 AED needs immediate service. Refer to “<i>Troubleshooting</i>” (Section 5.6 of this manual) or call Defibtech for service.</li> <li>• <b>No Flashing Light:</b> The DDU-2200 AED needs immediate service. Refer to “<i>Troubleshooting</i>” (Section 5.6 of this manual) or call Defibtech for service.</li> </ul> |
|---|--|




AED Status Screen



To check the status of the AED when the unit is off, press the **center softkey button**. The display screen will show unit status, battery pack status, and pad status. After a short period of time, the display screen, and the unit will turn off.

### 3.5 Installing the Defibtech Data Card (DDC Card) (optional)

The Defibtech Data Card (DDC card) is used to store event and audio information collected by the AED. All DDU-2200 AEDs will operate without DDC cards and will still store select event information internally. Information stored on the DDC card is retrievable with a separate Defibtech PC-based software package (refer to the “DefibView” section in Chapter 8 of this manual).



Before installing the DDC card, ensure the AED is turned OFF. Locate the data card/USB port access door on the right-hand side of the unit. Open the data card/USB port access door by slightly pushing and then sliding the door down to release the latch. The door will spring open. Insert the DDC card into the thin slot in the side of the AED centered above the USB port opening, notched end first, label side up, until it clicks into place. The card should be flush with the surface of the slot. If the card does not push in all the way, it may have been inserted upside down. In that case, remove the card, flip it over, and try inserting it again.

To remove the DDC card, press the card as far as it will go and then release. Upon release, the DDC card will be partially ejected and can be removed by pulling the DDC card the rest of the way out.

Close the data card/USB port access door by closing and then pushing the door up until the door latch engages.



Using non-Defibtech Data Cards (DDC cards) may damage the unit and will void the warranty.

### 3.6 Completing the Installation

Once you have completed the previous steps to set up your DDU-2200 AED, follow this procedure:

1. Turn the unit on by pressing the ON/OFF button.
2. Listen for the “Call for Help” voice prompt.
3. Turn the unit off by pressing and holding the ON/OFF button.
4. Listen for the “Powering Off” voice prompt.
5. Check Active Status Indicator (ASI) to verify that it is flashing green.

(Refer to the “Self-Tests” section in Chapter 5 of this manual for instructions on how to run a manually-initiated self-test.)

### 3.7 Storing the DDU-2200 AED

Store the DDU-2200 AED, with defibrillation pads attached, in environmental conditions within range of the specifications (refer to the “Environmental” section in Chapter 9 of this manual). The unit should also be stored so that the Active Status Indicator (ASI) can be readily seen.

The Active Status Indicator (ASI) should periodically flash with a green light. If it flashes with a red light or does not flash at all, the DDU-2200 AED needs servicing (refer to Section 3.4, “Checking the DDU-2200 AED Status,” for more information).

Defibtech recommends storing your AED in an easily accessible location where the unit can be seen and heard.




## 4 Using the DDU-2200 AED

This chapter describes how to use the DDU-2200. During a rescue, the unit analyzes the patient's rhythm and automatically charges if a shockable rhythm is detected. Push-button controls and indicators include an ON/OFF button, three softkey buttons, and a SHOCK Required LED indicator. The unit includes a display screen. Concise and easily understandable voice messages and text and video prompts guide the operator through the use of the unit.

The following sections describe in detail how to use the DDU-2200 AED. The basic steps for use are:


- Turn the DDU-2200 AED ON by pressing the **ON/OFF** button.
- Plug in pads connector into Pad Connector Socket on AED if not yet plugged in.
- Place defibrillation pads on patient (**follow instructions on pads package**).
- Follow voice and display prompts.

### 4.1 Overview



Using the DDU-2200 AED (continued)

### Unit Video Display Screen



**Battery Indicator** – The Battery Indicator indicates the *approximate* remaining battery capacity.

**Main Screen** – The Main Screen displays video instructions to guide the user during a rescue.

**Text Prompts** – The Text Prompt Area displays text prompts to guide the user during a rescue.

**Softkey Buttons** – The Softkey Buttons are located to the right of the display screen. If a softkey button is active, it will have a softkey icon displayed next to it. The softkey buttons are used to navigate menus or select actions.

**Rescue Breathing Options Softkey Icon** – When this icon is present on the screen (during a rescue), the user may press the corresponding softkey button to select CPR coaching with compressions only (no breathing) or CPR coaching with compressions and breathing.


**Information Softkey Icon** – When this icon is present on the screen, the user may press the corresponding softkey button for additional information with video instruction. The additional information is context dependent; topics include preparing the patient and performing CPR. To exit, press the softkey button again.

## 4.2 Preparation

### Checking the DDU-2200 AED Status

Visually check the Active Status Indicator (ASI). The ASI should flash green. The ASI flashes green to indicate ready for use status. The ASI flashes red, solid red, or is not lit at all to indicate that service is required.

The ASI is powered by the battery pack. If the battery pack has been completely discharged or is not installed in the unit, the active status indication will not be available. In this case, immediately replace the battery pack or reinsert it into the unit to restore active status indication.

|   |  |
|---|--|
|  <p><b>Active Status Indicator (ASI)</b></p> | <ul style="list-style-type: none"> <li>• <b>Flashing Green:</b> The DDU-2200 AED is OFF and ready for use.</li> <li>• <b>Solid Green:</b> The DDU-2200 AED is ON and ready for use.</li> <li>• <b>Flashing or Solid Red:</b> The DDU-2200 AED needs immediate service. Refer to “<i>Troubleshooting</i>” (Section 5.6 of this manual) or call Defibtech for service.</li> <li>• <b>No Flashing Light:</b> The DDU-2200 AED needs immediate service. Refer to “<i>Troubleshooting</i>” (Section 5.6 of this manual) or call Defibtech for service.</li> </ul> |
|---|--|

### Turning On the DDU-2200 AED

Press the green ON/OFF button to turn the DDU-2200 AED on. The unit will emit a “beep” and the display screen will turn on. The ASI indicator next to the ON/OFF button will illuminate green anytime the AED is on. (To turn the unit off, press AND HOLD the ON/OFF button for approximately two seconds; the unit will emit a “beep” and power off.)


Using the DDU-2200 AED (continued)

### Call for Help

Call professional emergency services for help. As soon as the AED is turned on, the unit will prompt the user to “Call for Help.” This is a reminder that the first step in a rescue should always be to contact professional emergency services.

If another person is available, the user should direct that person to call for help and then continue the rescue without delay.

### Selecting an Alternate Spoken Language




Some AED models are factory configured to support an alternate spoken language. If the AED supports an alternate language, a **Language Softkey Icon** (represented in the form of a flag) will be displayed. When the Language Softkey Icon is present, the user may press the corresponding softkey to switch the spoken voice prompts to the alternate language. **(Note: Text prompts continue in the primary language.)** Pressing the softkey again will switch the spoken language back to the primary language. The Language Softkey Icon is available only until the defibrillation pads are properly applied to the patient.

### Preparing the Patient

Prepare the patient by removing any clothing from the patient’s chest. Wipe away moisture from the chest if necessary (the defibrillation pads will stick better on dry skin). If necessary, shave excessive chest hair, which can prevent effective patient-pad contact. To ensure that defibrillation pads fully contact the patient’s skin, check that no jewelry or other objects are directly underneath where the pads will be placed.

### Connecting the Defibrillation Pads to the DDU-2200 AED



Connect pads to unit, if not already connected. Follow AED voice and display instructions. The DDU-2200 AED is designed to be stored with the defibrillation pads connector attached to the unit, while the pads themselves remain sealed in their package. This reduces the time needed to set up and start treatment in an emergency.

The Defibtech AED should be stored with the pads connector attached to the unit. However, if pads were damaged or not properly connected, you may need to substitute a new set of pads during an emergency. The pads connector socket is on the top left corner of the AED.

To detach a set of pads from the unit, pull firmly on the pad connector. Do not reuse used pads. Insert the connector for the new pads as shown above. The connector will only fit in one way – if the connector does not fit, rotate the connector before trying again. Insert connector firmly until it is completely seated in the unit.



When this **Information Softkey Icon** is present on the screen, the user may press the corresponding softkey button for additional information with video instruction. To exit, press the softkey button again.

### Opening the Defibrillation Pads Package

Remove the pads package from the pad storage slot at the back of the AED. Open the pads package by tearing along the dotted line, starting at the black arrow (follow directions on the package). Check that the pads are:

- Free from obvious signs of damage.
- Clean of excessive debris (for example, dirt if the pad was dropped).
- Not dried out, and that the gel is sticky and will adhere to the patient.
- Not expired. Do not use pads after the expiration date printed on the package.

If any of these conditions are found, use a new set of pads, if possible.

Using the DDU-2200 AED (continued)


### Applying the Defibrillation Pads to the Patient

Follow this procedure to apply the defibrillation pads to the patient:



1. Tear the defibrillation pads package along the dotted line near the top of the package.
2. Remove the pads from the package and follow the directions and diagram showing proper pad placement located on the pads package.
3. Peel off the protective backing from one of the pads before placing it as shown on the diagram on the pad. Peel the backing off only when the pad is ready to be placed.
4. Place the pad with the sticky side of the pad on the patient's skin.
5. Repeat steps 3 and 4 to place the other pad onto the patient.

Correct pad placement (shown below) is essential for effective analysis of the patient's cardiac rhythm and subsequent shock delivery (if required). Pad placement on infants and children under 8 years or less than 55 pounds (25 kg) is different than placement for adults and children 8 years or older or over 55 pounds (25 kg). If you are unsure of a child's age or weight or do not have child/infant pads, do not delay treatment.

Sample Adult Pads




Sample Child/Infant Pads



**For adults and children 8 years or older or over 55 pounds (25 kg), use adult pads:**

Place one pad just below the patient's right collar bone as shown in the picture. Place the second pad over the ribs on the patient's left side below the left breast.

Use picture on pad to determine individual pad placement.



**For infants and children under 8 years or less than 55 pounds (25 kg), use child/infant pads**

**(Note: Child/infant pads can be identified by their blue connector and pads package):**

Place one pad in the center of the chest and one pad on the center of the back, as shown. Use picture on pad to determine individual pad placement.

Using the DDU-2200 AED (continued)

### Follow DDU-2200 AED Instructions

At this point, the DDU-2200 AED will check to make sure that the pads are well connected to the patient and that an adequate ECG signal is being received. Do not touch the patient. Eliminate any patient movement, and cease CPR at this time.

If there is a problem with the pad connection, socket connection, patient motion, or other interference, the AED will guide the operator with both audible and displayed instructions. Text prompts are identical to or are an abbreviated form of the audio prompts. Video prompts reinforce the audio and text prompts and aid in high ambient noise environments.

## 4.3 Heart Rhythm Analysis

Once the DDU-2200 AED has determined that the pads are making a good connection to the patient, the AED will start the ECG rhythm analysis. The unit analyzes the ECG signal and determines whether a shockable or non-shockable rhythm is present. While analyzing, the AED will continue to monitor signal and pad conditions and will reassess analysis and prompt the user if additional user action is needed.

## 4.4 Delivering the Shock

If the fully-automatic DDU-2200 AED determines that a shock is advised, the unit will indicate that a shock is advised both via a vocal announcement and by flashing the Shock Required LED indicator for five seconds. While the AED charges, the unit continues to analyze the patient's heart rhythm as well as signal and pad conditions and may emit a charging tone. After completing charge-up, the unit will **automatically** deliver the shock. **Do not touch the patient while the Shock Required LED indicator is flashing.** If the unit detects that the heart rhythm has changed to non-shockable, the unit prompts the user to begin CPR.

The user can abort charging or shock delivery at any time by pressing and holding the ON/OFF button for approximately two seconds to turn the unit off.

## 4.5 CPR Period

The operator will be prompted to begin CPR, if needed. The unit will not monitor the patient's ECG rhythm during the CPR period. During the CPR period, the AED will not advise the user to "Stop Motion" even if motion is present. The user should follow the AED instructions during this time. Once the CPR period is complete, the unit will continue in Heart Rhythm Analysis mode.

CPR coaching is provided through a series of voice and visual prompts and audible tones. The factory default setting provides prompts for chest compressions only (no breathing).

However, breathing prompts can be enabled/disabled by pressing the softkey button next to the Rescue Breathing Options icon displayed on the screen during rescue (refer to the "*Rescue Breathing Options Softkey Icon*" section below). Breathing prompts can also be enabled/disabled by setting the menu option in Maintenance Mode (refer to "*CPR Breathing*" in Section 6.8 of this manual).



**Rescue Breathing Options Softkey Icon:** During a rescue, when this icon is present on the screen, the user may press the corresponding softkey button to select CPR coaching with compressions only (no breathing) or CPR coaching with compressions and breathing.

**Note:** Refer to "*CPR Breathing*" in Section 6.8 of this manual for instructions on how to change the factory default setting.



When this **Information Softkey Icon** is present on the screen, the user may press the corresponding softkey button for additional information with video instruction. To exit, press the softkey button again.

Using the DDU-2200 AED (continued)

## 4.6 Post-Use Procedures

After the DDU-2200 AED has been used on a patient, the unit should be cleaned following procedures in the "Cleaning" section in Chapter 5 of this manual and prepared for the next use. The following steps should be performed:

1. Connect a new pads package (check to make sure the package is not expired and package is not damaged).
2. Perform a self-test manually. Unit will report status at the end of the self-test (refer to the "Self-Tests" section in Chapter 5 of this manual for instructions on how to run a manually-initiated self-test).
3. Turn off the unit by pressing the ON/OFF button.
4. Check to make sure that the Active Status Indicator (ASI) is flashing green.

## 4.7 Voice and Text Prompts

The following section provides brief descriptions of some of the voice and text prompts that the user will hear and see when using the DDU-2200 AED.

### General Prompts

| <b>Voice</b>   | <b>Text</b>           |
|--|-----------------------|
| <b>"Call for help"</b>   | <b>Call for Help</b>  |
| <i>Purpose:</i> As soon as the DDU-2200 AED is turned on, the user will be prompted to call for help. This indicates that the first step in a rescue should always be to contact professional emergency services. If another person is available, the user should direct that person to call for help and then continue the rescue without delay.                                |                       |
| <b>"Pediatric mode"</b>  | <b>Pediatric Mode</b> |
| <i>Purpose:</i> This informs the user that child/infant pads are attached to the unit. Child/infant pads should only be used if the patient is an infant or a child under the age of 8 or less than 55 pounds (25 kg). For children 8 years or older or over 55 pounds (25 kg) and for adults, adult pads should be used. Do not delay therapy to determine exact age or weight. |                       |
| <b>"Training pads"</b>   | <b>Training Pads</b>  |
| <i>Purpose:</i> This informs the user that training pads are attached to the unit. Training pads are used for training purposes only and will not deliver a shock. In a rescue, immediately replace the training pads with defibrillation pads.  |                       |
| <b>"Powering off"</b>  | <b>Powering Off</b>   |
| <i>Purpose:</i> This informs the user that the unit is turning off.  |                       |

### Pad Connection/Pad Application-Related Prompts

| <b>Voice</b>  | <b>Text</b>                |
|---|----------------------------|
| <b>"Follow instructions to apply pads"</b>  | <b>Follow Instructions</b> |
| <i>Purpose:</i> This instructs the user to follow the AED prompts in order to apply the pads to the patient.  |                            |
| <b>"Remove clothing from patient's chest"</b>   | <b>Remove Clothing</b>     |
| <i>Purpose:</i> This instructs the user to remove all clothing from patient's chest. Pads must be applied to the patient's bare chest.  |                            |
| <b>"Locate pads package in back of AED"</b>   | <b>Locate Pads</b>         |
| <i>Purpose:</i> This helps the user locate the pads in the pad storage area, which is located on the back of the unit.  |                            |
| <b>"Plug in pads connector"</b>   | <b>Plug In Pads</b>        |
| <i>Purpose:</i> The DDU-2200 AED is unable to detect that the pads are plugged in. Check that the connector is fully inserted into the unit. If the pads are properly plugged in, continue to follow audio and visual instructions. |                            |
| <b>"Tear open pads package"</b>   | <b>Open Pads Package</b>   |
| <i>Purpose:</i> This instructs the user to tear open the pads package on the dotted line on the top of the package. Once the package is open, the user will be able to remove the pads from inside the package.                     |                            |

*Pad Connection/Pad Application-Related Prompts (continued)*

| <b>Voice</b>   | <b>Text</b>                                   |
|--|---|
| <b>"Peel adhesive pads from blue liner"</b>  | <b>Peel Pads</b>                              |
| <i>Purpose:</i> This instructs the user to peel each pad from the blue liner before placing the pads on the patient. Peel the pads from the blue liner only when the pad is ready to be placed. Place the pads with the sticky side of the pad on the patient's bare skin.   |   |
| <b>"Apply pads to patient's bare chest as shown"</b>   | <b>Apply Pads to Patient</b>                  |
| <i>Purpose:</i> The DDU-2200 AED has determined that the pads are not placed on the patient or not properly applied. Place pads on the patient following instructions on the pads package. If the prompts continue, try replacing the pads with a new set.   |   |
| <b>"Poor pad contact to patient"<br/>"Press pads firmly"</b>   | <b>Poor Pad Contact<br/>Press Pads Firmly</b> |
| <i>Purpose:</i> The pads are not making proper contact with the patient and the impedance is out of range for proper ECG analysis and shock delivery. Check that the pads are properly placed and fully adhering to the patient and that there are no air bubbles between the pads and the patient. If the pads are not sticking due to moisture, dry the patient. If the pads are not sticking due to excessive hair, shave or clip excessive chest hair. If the prompts continue, try replacing the pads with a new set. |   |
| <b>"Check pads"</b>  | <b>Check Pads</b>                             |
| <i>Purpose:</i> The pads are making improper contact with the patient or touching each other and the impedance is out of range for proper ECG analysis and shock delivery. Check that the pads are not touching each other and that the patient is dry. If the prompts continue, try replacing the pads with a new set.  |   |
| <b>"Pausing for CPR"</b>   | <b>Pausing for CPR</b>                        |
| <i>Purpose:</i> If too long a period of time has passed, the user should stop attempting to resolve problems with the pads and assess the condition of the patient. The user will be prompted to begin CPR.  |   |
| <b>"Replace pads"</b>  | <b>Replace Pads</b>                           |
| <i>Purpose:</i> If another set of pads is available, replace the pads. Otherwise, check that the pads are properly placed and fully adhered to the patient. Make sure that the pads are not touching each other. If the pads are not sticking due to moisture, dry the patient. If the pads are not sticking due to excessive hair, shave or clip excessive chest hair.  |   |

**Motion/Interference Prompts**

| <b>Voice</b>  | <b>Text</b>              |
|---|--------------------------|
| <b>"Stop motion"</b>  | <b>Stop Motion</b>       |
| <i>Purpose:</i> The DDU-2200 AED has detected possible motion in the patient. Stop all patient motion, including CPR, in response to this prompt.   |                          |
| <b>"Stop interference"</b>  | <b>Stop Interference</b> |
| <i>Purpose:</i> The DDU-2200 AED has detected interference on the ECG signal. Eliminate any radio or electrical sources of interference. Check the pads to make sure they are adhering properly to the patient. If the environment is very dry, minimize movement around the patient to reduce static discharges. |                          |
| <b>"Pausing for CPR"</b>  | <b>Pausing for CPR</b>   |
| <i>Purpose:</i> The user should stop attempting to resolve motion and/or interference problems and assess the condition of the patient. The user will be prompted to begin CPR.   |                          |

## Heart Rhythm Analysis Prompts

| <b>Voice</b>   | <b>Text</b>                                 |
|--|---|
| <b>"Analyzing heart rhythm"</b><br><b>"Analyzing"</b>  | <b>Analyzing Rhythm</b><br><b>Analyzing</b> |
| <i>Purpose:</i> The DDU-2200 AED is actively analyzing the patient's ECG signal. The AED will continue analyzing until it has determined whether a rhythm is shockable or non-shockable or if analyzing is interrupted for some reason.  |   |
| <b>"Do not touch the patient"</b>  | <b>Do Not Touch Patient</b>                 |
| <i>Purpose:</i> The DDU-2200 AED is trying to analyze the patient's heart rhythm. The operator should not touch the patient. This prompt will be spoken at the beginning of the analysis period.   |   |
| <b>"Analyzing interrupted"</b>   | <b>Analyzing Interrupted</b>                |
| <i>Purpose:</i> The DDU-2200 AED has determined that accurate ECG analysis is not possible and has ceased analyzing. The operator is prompted to resolve the problem (refer to the "Motion/Interference Prompts" and the "Pad Connection/Pad Application-Related Prompts" section in this chapter). Once the problem is resolved, the unit will enter analysis mode again. |   |
| <b>"No shock advised"</b>  | <b>No Shock Advised</b>                     |
| <i>Purpose:</i> The DDU-2200 AED has determined that a shock is not required. The unit will <b>NOT</b> charge and deliver a shock. The user will be prompted to begin CPR.   |   |
| <b>"Shock advised"</b>   | <b>Shock Advised</b>                        |
| <i>Purpose:</i> The DDU-2200 AED has determined that a shock is recommended and the unit will begin charging in anticipation of delivering a defibrillation shock.   |   |

## Shock-Related Prompts

| <b>Voice</b>  | <b>Text</b>                |
|---|----------------------------|
| <b>"Charging"</b>   | <b>Charging</b>            |
| <i>Purpose:</i> The DDU-2200 AED has determined that a shock is recommended and is charging the unit in anticipation of a defibrillation shock. Analysis will continue during this phase. A tone may sound to indicate charging progress. If the unit detects a rhythm change to a non-shockable one, charging will abort and the user will be prompted to begin CPR.<br><b>Note:</b> At any time during the charging process or after the AED has been charged, the operator may disarm the unit by pressing the ON/OFF button for approximately 2 seconds to power off the AED. |                            |
| <b>"Stand clear"</b>  | <b>Stand Clear</b>         |
| <i>Purpose:</i> The DDU-2200 AED is charging and the operator and others should stand clear of the patient. Analysis will continue during this phase and analyzing prompts will continue to be displayed. A tone may sound to indicate charging progress. If the unit detects a rhythm change to a non-shockable one, charging will abort and the user will be prompted to begin CPR.   |                            |
| <b>"Shocking in 3, 2, 1"</b>  | <b>Delivering Shock</b>    |
| <i>Purpose:</i> This indicates that the DDU-2200 AED has fully charged, that the heart rhythm analysis algorithm still indicates a shock is recommended, and the unit is about to deliver a shock. The SHOCK Required Indicator will flash during this time. The shock will automatically be delivered after the count reaches "1." <b>Do not touch the patient during this time.</b>   |                            |
| <b>"Shock 'x' delivered"</b>  | <b>Shock "x" Delivered</b> |
| <i>Purpose:</i> The DDU-2200 AED has delivered the shock. The 'x' indicates the number of shocks that have been delivered since the unit was turned on. After each shock, the AED will enter Post-Shock CPR mode.   |                            |
| <b>"Shock cancelled"</b>  | <b>Shock Cancelled</b>     |
| <i>Purpose:</i> The DDU-2200 AED has aborted the shock. If the unit detects a rhythm change to a non-shockable rhythm, the unit will cancel the shock.  |                            |



## No Shock Required Prompts

| <b>Voice</b>  | <b>Text</b>   |
|---|---|
| <b>"No shock advised"</b><br><b>"It is safe to touch the patient"</b>   | <b>No Shock Advised</b><br><b>OK to Touch Patient</b> |
| <i>Purpose:</i> The DDU-2200 AED has determined that a shock is not required. The unit will not charge or automatically deliver the shock. If the AED is charged, the shock will be canceled. The user will be prompted to begin CPR. |   |

## CPR Prompts

**Note:** CPR breathing coaching prompts can be set through the **Rescue options** menu option listed on the **AED Main Menu** screen. The factory default setting provides prompts for chest compressions only (no breathing). Breathing prompts can be included either by changing the menu option (refer to "CPR Breathing" in Section 6.8 of this manual) or by pressing a softkey button during rescue (refer to "Rescue Breathing Options Softkey Icon" in Section 4.5).

| <b>Voice</b>   | <b>Text</b>                        |
|--|------------------------------------|
| <b>"Begin CPR now"</b>   | <b>Begin CPR Now</b>               |
| <i>Purpose:</i> This indicates that the user should begin performing CPR immediately. The unit will not monitor the patient's ECG rhythm during this CPR period.   |                                    |
| <b>"Give Compressions"</b>   | <b>Give "xx" Compressions</b>      |
| <i>Purpose:</i> This indicates that the user should begin CPR compressions immediately. The unit will emit a beep at the rate that compressions should be given.   |                                    |
| <b>"Continue"</b><br><b>"Continue for 1 minute 'xx' seconds"</b>   | <b>Continue "xx" Seconds</b>       |
| <i>Purpose:</i> This indicates that the user should continue performing CPR. This phrase is spoken to let the user know that the unit is still operating normally. The unit will not be monitoring the patient's ECG rhythm during this mandatory two minute CPR period. |                                    |
| <b>"Ending in 5, 4, 3, 2, 1"</b>   | <b>Ending in "xx" Seconds</b>      |
| <i>Purpose:</i> This indicates that the user should prepare to finish performing CPR. This phrase is spoken during the last several seconds of the CPR period to let the operator know that the unit is still operating normally and that the CPR period is ending.      |                                    |
| <b>"Stop CPR"</b><br><b>"Stop Now"</b>   | <b>Stop CPR</b><br><b>Stop Now</b> |
| <i>Purpose:</i> This indicates that the CPR period has ended and the user should stop CPR.   |                                    |

## CPR Coaching Help Prompts

| <b>Voice</b>  | <b>Text</b>   |
|---|---|
| <b>"Place hands"</b>  | <b>Place Hands</b>  |
| <i>Purpose:</i> This reminds the user of the correct placement of hands for CPR.  |   |
| <b>"Press"</b><br><b>"Compress Chest"</b>   | <b>Press</b><br><b>Compress Chest</b>                                 |
| <i>Purpose:</i> This reminds the user to perform CPR compressions.  |   |
| <b>"Tilt head back"</b><br><b>"Pinch nose"</b><br><b>"Give rescue breaths"</b>  | <b>Tilt Head Back</b><br><b>Pinch Nose</b><br><b>Give "x" Breaths</b> |
| <i>Purpose:</i> This guides the user to prepare the patient for rescue breaths and to give breaths.   |   |
| <b>"Breathe"</b>  | <b>Breathe</b>  |
| <i>Purpose:</i> This instructs the user to give rescue breaths. Each time the instruction is given, the user should give the patient a rescue breath. |   |

Using the DDU-2200 AED (continued)

## 4.8 Operational Environment

---

The Defibtech DDU-2200 AED is designed to operate in a wide range of environmental conditions. To ensure the reliability and safety of the AED in a given environment, refer to the “*Environmental*” section in Chapter 9 of this manual for a detailed list of approved environmental conditions.

# 5 Maintenance and Troubleshooting

This chapter describes the maintenance and troubleshooting procedures for the DDU-2200 AED. The self-tests that are automatically performed by the device are described along with recommended routine maintenance. A troubleshooting guide is provided to help diagnose user serviceable problems.

The DDU-2200 AED contains no user serviceable parts.

## 5.1 Routine Unit Maintenance

The DDU-2200 AED is designed to be very low maintenance. Simple maintenance tasks are recommended to be performed regularly to ensure its readiness (see sample maintenance table below). Different maintenance intervals may be appropriate depending on the environment where the AED is deployed, and ultimately the maintenance program is at the discretion of the emergency response program's medical director.

| Daily | Monthly | After Each Use | Action   |
|-------|---------|----------------|--|
| •     | •       | •              | Check that the Active Status Indicator (ASI) is flashing green |
|       | •       | •              | Check the condition of the unit and accessories                |
|       |         | •              | Run manually-initiated self-test                               |
|       |         | •              | Replace pads   |
|       | •       |                | Check pad and battery pack expiration dates                    |
|       |         | •              | Check the DDC card, if one was installed                       |


**Note:** If the unit has been dropped, mishandled, or abused, a manually-initiated self-test should be performed.

### Checking the Active Status Indicator

The Active Status Indicator (ASI) is located to the right of the ON/OFF button of the DDU-2200 AED and indicates the operational readiness state of the unit. The ASI will periodically flash green to indicate that the unit is ready for use. Ready for use means that the DDU-2200 has passed the most recent self test (scheduled or user initiated). If it is flashing red, solid red, or not flashing at all, the AED needs service. Anytime the ASI is flashing red, the unit will periodically emit two "beeps" to call attention to it.

If the ASI is not flashing at all, the most likely cause is that the battery pack needs to be replaced (refer to the "Installing and Removing the Battery Pack" section in Chapter 3 of this manual). Once the battery pack has been replaced with a fresh battery pack, the ASI should once again flash green. If it still does not flash green after inserting a new battery pack, the DDU-2200 AED is non-operational and may need servicing. Call Defibtech for service (refer to the "Contacts" section in Chapter 12 of this manual).


If the ASI is flashing red, turn the DDU-2200 AED on. If the unit does not turn on or does not speak, the AED is non-operational and requires servicing. If the unit does turn on, then turn the unit off and the voice prompts will indicate the nature of the problem.

|   |  |
|---|--|
|  | <ul style="list-style-type: none"> <li>• <b>Flashing Green:</b> The DDU-2200 AED is OFF and ready for use.</li> <li>• <b>Solid Green:</b> The DDU-2200 AED is ON and ready for use.</li> <li>• <b>Flashing or Solid Red:</b> The DDU-2200 AED needs immediate service. Refer to "Troubleshooting" (Section 5.6 of this manual) or call Defibtech for service.</li> <li>• <b>No Flashing Light:</b> The DDU-2200 AED needs immediate service. Refer to "Troubleshooting" (Section 5.6 of this manual) or call Defibtech for service.</li> </ul> |
|---|--|


**Active Status Indicator (ASI)**

### Checking the AED Status Using the AED Status Screen

You may also check the status of the unit when it is off by pressing the center softkey button to enter Maintenance Mode and display the AED Status Screen.

|  |  |
|--|--|
|  <p>The AED Status Screen shown at left is used to provide a quick overview of the DDU-2200 AED's status and to display select information without turning the unit on in Rescue Mode.</p> <p>With the AED off, press and release the <b>CENTER</b> softkey button to display the AED Status Screen. The AED Status Screen will be displayed for a short period of time.</p> <p>If the unit does not turn on, check to make sure a good battery pack is installed (refer to the "Troubleshooting" section in this chapter).</p> <p>From the AED Status Screen you can enter Maintenance Mode (see Chapter 6 for details) by pressing the softkey button to the right of the <b>Tool Icon</b>.</p> | <p>The AED Status Screen shown at left is used to provide a quick overview of the DDU-2200 AED's status and to display select information without turning the unit on in Rescue Mode.</p> <p>With the AED off, press and release the <b>CENTER</b> softkey button to display the AED Status Screen. The AED Status Screen will be displayed for a short period of time.</p> <p>If the unit does not turn on, check to make sure a good battery pack is installed (refer to the "Troubleshooting" section in this chapter).</p> <p>From the AED Status Screen you can enter Maintenance Mode (see Chapter 6 for details) by pressing the softkey button to the right of the <b>Tool Icon</b>.</p> |
|--|--|

**Note:** If the unit requires service, the AED Status Screen will present information about the problem to the user. The user should follow the text prompts to address the condition requiring attention.

|   |  |
|---|--|
|  | <p><b>Card Application Softkey Icon:</b> If an application is present on an inserted Defibtech Data Card (DDC card), a card icon will also appear next to the center softkey button. Pressing this button will load and run the application contained on the card.</p> |
|---|--|

### Maintenance-Related Prompts

| Voice  | Text  |
|--|---|
| <p><b>"Power on test failed"</b><br/><b>"Service code 'xxxx' "</b></p>   | <p><b>Power On Test Failed</b><br/><b>Service Code "xxxx"</b></p> |
| <p><i>Purpose:</i> This indicates that the DDU-2200 AED has failed the power-on self-test and may be non-operational and may require servicing. The code number will indicate to service personnel the type of problem the unit is experiencing.</p> |   |
| <p><b>"Battery test failed"</b><br/><b>"Service code 'xxxx' "</b></p>  | <p><b>Battery Test Failed</b><br/><b>Service Code "xxxx"</b></p>  |
| <p><i>Purpose:</i> This indicates that the DDU-2200 AED's battery pack is non-operational and needs servicing. The code number will indicate to the service personnel the type of problem that the unit is experiencing.</p>                         |   |
| <p><b>"Service code 'xxxx' "</b></p>   | <p><b>Service Code "xxxx"</b></p>                                 |
| <p><i>Purpose:</i> The DDU-2200 AED will report this message when it powers off, indicating a service code that was previously detected.</p>   |   |
| <p><b>"Service required"</b></p>   | <p><b>Service Required</b></p>                                    |
| <p><i>Purpose:</i> This indicates that the DDU-2200 AED has detected an internal error, is non-operational, and needs servicing.</p>   |   |
| <p><b>"Battery low"</b></p>  | <p><b>Battery Low</b></p>   |
| <p><i>Purpose:</i> This indicates that the battery pack capacity is low and that the battery pack should be replaced soon. The AED will still be able to deliver at least three defibrillation shocks the first time this message is spoken.</p>     |   |
| <p><b>"Replace battery now"</b></p>  | <p><b>Replace Battery Now</b></p>                                 |
| <p><i>Purpose:</i> This indicates that the battery pack is almost discharged and that the AED may not be able to deliver defibrillation shocks. Replace the battery pack immediately.</p>  |   |
| <p><b>"Pads missing"</b></p>   | <p><b>Pads Missing</b></p>  |
| <p><i>Purpose:</i> This indicates that the unit did not detect connected pads during a self-test.</p>  |   |
| <p><b>"Pads expired"</b></p>   | <p><b>Pads Expired</b></p>  |
| <p><i>Purpose:</i> This indicates that defibrillation pads have expired. Replace the pads immediately.</p>   |   |

## Checking the Condition of the Unit and Accessories

Inspect the unit for dirt and contamination, especially in the pads connector socket and around the battery pack opening (refer to Section 5.3 for guidance on cleaning your AED).

Inspect the unit display screen for damage. Look for cracks or other signs of damage on the case, especially near the connector socket.

If any cracks or other signs of damage are visible, remove the AED from service and contact an authorized service center.


## Replacing Pads

The Defibtech defibrillation pads are intended for one-time use only. The pads must be replaced after each use or if the package has been damaged.

The DDU-2200 AED defibrillation pads are supplied in a sealed pouch with the connector and part of the cable exposed. The DDU-2200 AED is designed to be stored with the electrode cable already installed. This allows the pads to be stored in a pre-connected state for rapid deployment during an emergency.



**DO NOT** remove the defibrillation pads from the sealed package until the pads are to be used. The packaging should be opened only immediately prior to use, otherwise the pads may dry out and become non-functional.



**STEP 1: Inspect the Pads** – First, check to ensure that the pads package has not expired. Do not use pads past their expiration date. Discard expired pads. Next, check to ensure that the pads package has not been torn, opened, or damaged. Dispose of the pads if the package is open or damaged. Inspect the pads cable and replace pads if any nicks, cuts, or broken cables are found.

**STEP 2: Connect the Pads to the Unit** – Insert the connector end of the defibrillation pad cable into the pads connector socket on the top-left corner of the DDU-2200 AED as shown. Press the pads connector in firmly until it is fully seated in the unit.

**STEP 3: Store the Pads in Back of Unit** – The pads package can then be stored in the pad storage slot in the back of the DDU-2200 AED. After connecting the pads connector to the unit, push the pads package, rounded end first, with the pictures on the package facing up and out, into the pad holder compartment on the back of the AED. When the pads package is fully inserted, press the pad cable into the groove in the back of the unit to hold it in place and tuck any excess cable behind the pads package.



The defibrillation pads are intended for one-time use only and must be discarded after use or if the package has been opened.

### Checking Pad and Battery Pack Expiration Dates


It is important to check the expiration dates of the pads and battery pack. Both may be checked when the unit is off by pressing the center softkey button to display the AED Status Screen (refer to the “AED Status Screen” section in Chapter 6 of this manual).

The expiration dates can also be found on each item: the pads package expiration date is printed on the outside of the sealed package; the battery pack expiration date is printed on the label on the battery pack.

Once an accessory is past its expiration date, it should be taken out of service and replaced as soon as possible. Follow the instructions in the “Installing and Removing the Battery Pack” and “Connecting the Defibrillation Pads” sections of Chapter 3 in this manual to replace an expired part with an unexpired part. Pads should be discarded. Battery packs should be appropriately recycled.

### Checking the Defibtech Data Card (DDC Card)

Each time the DDU-2200 AED is used, an event file is created on the Defibtech Data Card (DDC card), if installed. If the unit was used to treat a patient, the DDC card in the unit should be removed and provided to the patient’s care provider. A new DDC card should be installed before the next use.



To remove the DDC card, ensure the AED is OFF. Locate the data card/USB port access door on the right-hand side of the unit. Open the data card/USB port access door by slightly pushing and then sliding the door down to release the latch. The door will spring open. To remove the DDC card, press the card in as far as it will go, and then release. Upon release, the DDC card will be partially ejected and can be removed by pulling the DDC card the rest of the way out.

To install a new DDC card, insert the DDC card into the thin slot in the side of the AED centered above the USB port opening, notched end first, label side up, until it clicks into place. The card should be flush with the surface of the slot. If the card does not push in all the way, it may have been inserted upside down. In that case, remove the card, flip it over, and try inserting it again.

Close the data card/USB port access door by closing and then pushing up the door until the door latch engages.

**Note:** A DDC card is not required for the DDU-2200 AED to operate. Even if a DDC card is not installed, the unit will still record basic essential information internally. The AED will still operate properly even after a “replace data card” message.

## 5.2 Self-Tests

---

The DDU-2200 AED provides for both automatic and manually-initiated self-tests. These self-tests test various components of the AED, including the system controls, battery pack condition, charge/discharge functions, and measurement and signal acquisition functions.

### Automatic Unit Self-Tests

Every time the unit is turned on, power-on self-tests are performed to test the basic operation of the unit. The unit also performs daily, weekly, monthly, and quarterly self-tests automatically (without any intervention from the operator) to check the integrity of the unit's hardware and software. The unit will also perform a battery insert self-test when the battery pack is inserted.

### Manual Self-Tests

Manually-initiated self-tests may be run at any time by the user to test the DDU-2200 AED's systems, including the charging and shocking functions (the shock is internally dissipated and no voltage will be present at the pads).

To run a manually-initiated AED test, the unit must be put into Maintenance Mode (refer to the "AED Maintenance Screen" section in Chapter 6 of this manual for detailed information on performing these self-tests).

**Note:** Running a manually-initiated self-test will consume approximately one shock's worth of energy from the battery.

## 5.3 Cleaning

---

After each use, clean the DDU-2200 AED of any dirt or contaminants on the case and connector socket. The following are important guidelines to be adhered to when cleaning the device:

- The battery pack should be installed when cleaning the DDU-2200 AED.
- Do not immerse the DDU-2200 AED in fluids or allow fluids to enter the unit.
- Do not spray cleaning solutions directly on the unit or its connectors.
- Do not use abrasive materials or strong solvents such as acetone or acetone based cleaning agents.
- To wipe the DDU-2200 AED's case clean, use a soft cloth dampened with one of the following recommended cleaning agents:
  - Soapy water
  - Ammonia based cleaners
  - Hydrogen peroxide
  - Isopropyl alcohol (70 percent solution)
  - Chlorine bleach (30 ml/liter water)
- Ensure that the connector socket is completely dry before reinstalling the pads cable. After cleaning, allow the unit to completely dry. Before returning it to service, always check the AED operational status (refer to "Checking the AED Status Using the AED Status Screen" earlier in this chapter).

Please note that none of the items provided with the DDU-2200 AED (including the AED itself) are sterile or require sterilization.



Do not sterilize the DDU-2200 AED or its accessories.

## 5.4 Storage

---

The DDU-2200 AED should be placed in a readily accessible location in an orientation where the Active Status Indicator (ASI) next to the ON/OFF button can be easily seen and heard. In general, the unit should be stored in clean, dry and moderate temperature conditions. Make sure that the environmental conditions of the storage location are within the ranges detailed in the "Environmental" section in Chapter 9 of this manual.

## 5.5 Operator's Checklist

The following checklist may be used as the basis for an Operator's Checklist. The table should be copied and filled out as recommended by the schedule in the "Routine Unit Maintenance" section of this chapter. As each item is completed it should be checked off.

| <b>Defibtech DDU-2200 AED Operator's Checklist</b>   |  |  |  |  |  |  |
|--|--|--|--|--|--|--|
| Defibtech DDU-2200 AED Serial Number: _____  |  |  |  |  |  |  |
| Defibtech DDU-2200 AED Location: _____   |  |  |  |  |  |  |
| <b>Date:</b>   |  |  |  |  |  |  |
| Check unit and accessories for damage, dirt, and contamination. Clean or replace as necessary. |  |  |  |  |  |  |
| Check that spare battery pack and pads are available.  |  |  |  |  |  |  |
| Check that battery pack and pads are not past expiration dates.                                |  |  |  |  |  |  |
| Check that the Active Status Indicator (ASI) is flashing green.                                |  |  |  |  |  |  |
| Comments:  |  |  |  |  |  |  |
| Inspection by: (initials or signature)   |  |  |  |  |  |  |



## 5.6 Troubleshooting

The following table lists the symptoms, the possible causes, and the possible corrective actions for common problems. Refer to the other sections of this user manual for detailed explanations on how to implement the corrective actions. If the unit continues to be non-functional, refer the unit for servicing (see Chapter 12 of this manual for contact information).

| <b>Symptom</b>  | <b>Possible Cause</b>                             | <b>Corrective Action</b>  |
|---|---|---|
| <b>Unit will not turn on</b>  | Battery pack not inserted                         | Insert battery pack   |
|   | Battery pack depleted or needs servicing          | Replace battery pack or call for service  |
|   | Unit needs servicing                              | Call for service  |
| <b>Unit immediately turns off</b>                                       | Battery pack depleted                             | Replace battery pack  |
|   | Unit needs servicing                              | Call for service  |
| <b>ASI flashes red and/or unit makes periodic "beep" sound</b>          | Unit may need servicing                           | Go to AED Status Screen by pressing the CENTER softkey button or call for service |
|   | Battery pack non-functional                       | Replace battery pack  |
|   | Defibrillation pads are not pre-connected to unit | Connect defibrillation pads to unit   |
|   | Defibrillation pads or battery pack expired       | Replace expired component   |
| <b>ASI does not flash at all while unit is in standby (powered off)</b> | Battery pack not inserted                         | Insert battery pack   |
|   | Battery pack is low or needs servicing            | Replace battery pack or call for service  |
|   | Unit needs servicing                              | Call for service  |
| <b>"Power on test failed, service code 'xxxx'" prompt</b>               | Unit needs servicing                              | Record code number and call for service   |
| <b>"Battery test failed, service code 'xxxx'" prompt</b>                | Battery pack needs servicing                      | Record code number and call for service   |
| <b>"Service required" prompt</b>  | Unit needs servicing                              | Call for service  |
| <b>"Replace battery now" prompt</b>                                     | Battery pack capacity is critically low           | Unit may not deliver a shock, replace battery pack immediately                    |
| <b>"Battery low" prompt</b>   | Battery pack capacity is getting low              | Replace battery pack as soon as possible  |
| <b>Display screen does not work</b>                                     | Battery pack depleted                             | Replace battery pack  |
|   | Battery pack not inserted properly                | Make sure battery pack is oriented correctly and fully inserted                   |
|   | Unit needs servicing                              | Call for service  |
| <b>"Pads missing" prompt</b>  | Pads not connected to unit                        | Make sure pads connector is oriented correctly and fully inserted into unit       |
| <b>"Plug in pads connector" prompt</b>                                  | Pads connector not plugged in                     | Plug in pads connector  |
|   | Pads connector broken                             | Replace pads  |
|   | Unit's connector broken                           | Call for service  |

| <b>Symptom</b>   | <b>Possible Cause</b>  | <b>Corrective Action</b>                       |
|--|--|--|
| <b>"Pads expired" prompt</b>   | Pads are beyond expiration date shown on pads package              | Replace pads                                   |
| <b>"Apply pads to patient's bare chest as shown" prompt</b>  | Pads not connected to patient                                      | Place pads on patient                          |
|  | Pads not making good connection to patient                         | Check pad connection to patient                |
|  | Pads or pad cable damaged  | Replace pads                                   |
| <b>"Poor pad contact to patient", "Press pads firmly", "Replace pads", "Non-rescue pads" or "Warning" prompt</b> | Dry pads   | Replace pads                                   |
|  | Partial pad connection   | Check that pads are placed securely on patient |
|  | Pads touching  | Separate pads and place correctly on patient   |
|  | Non-rescue pads (e.g. trainer pads) connected while in Rescue Mode | Replace non-rescue pads with rescue pads       |
| <b>"Check pads" prompt</b>   | Pads touching  | Separate pads and place correctly on patient   |
| <b>"Stop motion" prompt</b>  | Patient motion has been detected                                   | Stop patient motion                            |
| <b>"Stop interference" prompt</b>  | External interference has been detected                            | Stop external interference                     |
| <b>"Analyzing interrupted" prompt</b>  | Motion or interference detected                                    | Stop motion or interference                    |
| <b>"Shock cancelled" prompt</b>  | Patient's ECG rhythm changed                                       | No action necessary                            |
|  | Low battery – insufficient to charge                               | Replace battery pack                           |
|  | Bad pad to patient connection                                      | Check that pads are placed securely on patient |
|  | Dry pads   | Replace pads                                   |
| <b>"Replace data card" prompt</b>  | DDC card is full   | Replace DDC card with a card that is not full  |
|  | DDC card has failed  | Replace DDC card                               |
| <b>"Replace data card" prompt</b>  | DDC card is full   | Replace DDC card with a card that is not full  |
|  | DDC card has failed  | Replace DDC card                               |

## 5.7 Repair

The DDU-2200 AED contains no user serviceable parts. If the unit needs servicing, call Defibtech (refer to Chapter 12 of this manual for contact information).

# 6 Maintenance Mode


## 6.1 Overview

Maintenance Mode for the Defibtech DDU-2200 AED permits the user to perform maintenance-related actions such as viewing unit information, initiating unit self-tests, changing unit parameters, downloading rescue data, and upgrading software.

Maintenance Mode is navigated through a series of screens, menus, and menu options. In Maintenance Mode, the softkey buttons located directly to the right of the display screen are used to scroll through and select menu options. When a softkey icon (for example, an arrow) appears on the display screen directly to the left of a softkey button, the softkey button is functional for that screen. If a softkey icon is not displayed on the screen, then the corresponding softkey button has no functionality for that screen.

**Note:** While the unit is in Maintenance Mode, it cannot perform a rescue. Maintenance Mode allows the user to go directly to Rescue Mode by selecting the **Rescue now** option. The **Rescue now** option appears at the top of every screen/menu when the unit is in Maintenance Mode. The user can also exit Maintenance Mode at any time and go to Rescue Mode by pressing the ON/OFF button to turn the unit off and then immediately pressing the ON/OFF button again to turn the unit back on.

### The Display Screen (During Maintenance Mode):



## 6.2 Navigation (in Maintenance Mode)

The three softkey buttons located to the right of the display screen are used to navigate in Maintenance Mode. Typical functions of the softkey buttons are the following:

- Top softkey button: Scroll up
- Center softkey button: Select highlighted option
- Bottom softkey button: Scroll down

When a menu option is highlighted and then selected (typically by pressing the CENTER softkey button), either another screen will be displayed with additional menu options or an action will be performed.


### Exiting Maintenance Mode

To exit Maintenance Mode and return to Rescue Mode, scroll to and select **Rescue now** or simply turn the unit off and then back on.

To exit Maintenance Mode and turn the unit off, scroll to and select **Turn AED off** or simply turn the unit off by pressing the ON/OFF button.

## 6.3 Entering Maintenance Mode

**Before You Begin:** Make certain the DDU-2200 AED is turned off and a battery pack is installed.



**STEP 1** – Press and release the CENTER softkey button.

**Result** – The unit will turn on and display the AED Status Screen for a short period of time.


If the unit does not turn on, check to make sure a good battery pack is installed (refer to the “*Troubleshooting*” section in Chapter 5 of this manual).

**STEP 2** – Press the **BOTTOM** softkey button (to the right of the Tool Icon). **Note:** If the bottom softkey button is not pressed within a short period of time, the unit will automatically turn off.

**Result** – The unit will enter Maintenance Mode and display the AED Main Menu screen.

### 6.4 AED Main Menu Screen

The AED Main Menu screen allows the user to view the status of the AED, perform maintenance functions, change AED options, and access help topics. All maintenance functions are accessed through the AED Main Menu screen. The user may select from the following options using the softkey buttons:



**Rescue now** – Puts device in Rescue Mode

**AED status** – Displays current AED status information

**AED maintenance** – Displays AED Maintenance Menu screen

**AED options** – Displays AED Options Menu screen

**Rescue options** – Displays Rescue Options Menu screen

**Help topics** – Displays Help Topics screen

**Turn AED off** – Turns unit off

When the user selects “Rescue now,” the unit will exit Maintenance Mode and proceed directly to Rescue Mode. The other menu options will perform different functions and are described in detail below.

### 6.5 AED Status Screen

The AED status screen displays unit-specific data such as current status, battery pack charge state, battery pack expiration date, defibrillation pads expiration date, unit serial number, battery pack serial number, and software version number.

**Before you begin:** Be sure the unit is in Maintenance Mode.

**To enter:** Navigate to **AED status**:

**AED Main Menu → AED status**

**Note:** When the unit is turned off, the AED status screen can also be accessed by pressing the center softkey button.

**What it does:** The unit will display the AED status screen. This is an informational screen only; no action is taken by the AED.

**To exit:** To exit the AED status screen, press and release the BOTTOM softkey button. The unit will exit the AED status screen and return to the AED Main Menu screen.

## 6.6 AED Maintenance Screen

The AED maintenance screen allows the user to select such options as AED tests, software upgrades, data backups, and data card functions.

**Before you begin:** Be sure the unit is in Maintenance Mode.

**To enter:** Navigate to **AED maintenance**:

**AED Main Menu → AED maintenance**

**What it does:** The unit will display the AED maintenance menu screen. This screen will allow the user to navigate further to perform various maintenance tasks:

- **Perform AED test**
- **Upgrade AED**
- **Transfer data to card**
- **Format data card**
- **Run application from card**

**To exit:** Use the TOP or BOTTOM softkey buttons to scroll to and highlight the selection “Go to main menu.” Press the CENTER softkey button. The unit will exit the AED maintenance screen and return to the AED Main Menu screen.

### ➔ Perform AED Test

**Perform AED test** will initiate a system hardware and software self-test.

**Note:** Running manually-initiated AED tests will consume approximately one shock’s worth of battery life.

**Before you begin:** Be sure the unit is in Maintenance Mode.

**To enter:** Navigate to **Perform AED test**:

**AED Main Menu → AED maintenance → Perform AED test**

**What it does:** When the user selects “Perform AED test” selection and presses the CENTER softkey button, the unit will begin performing the self-test procedure:

The unit says: “Performing AED test”

The unit displays: Testing AED

Follow the directions until the test is complete. Once the AED test is complete the unit will verbally and visually report the status of the AED. The information will be displayed in a pop-up window. The user must press any softkey button to confirm the test status and return to the AED Maintenance screen.

If self-test passes: The unit will speak and display: “AED OK”

If self-test fails: The unit will display an error screen with text prompts providing instructions to address the condition.

**Note:** If self-test fails, the user should follow the text prompts to address the condition requiring attention or refer to the “*Troubleshooting*” section in Chapter 5 of this manual.

**To exit:** Press any softkey button. The self-test status pop-up window will clear and the display will return to the AED Maintenance menu screen.

### ➔ Upgrade AED

The **Upgrade AED** menu selection is a method of performing a unit upgrade and will activate the system upgrade procedure from a Defibtech Data Card (DDC card) containing an upgrade application.

**Note:** Upgrades can also be performed directly from the AED Status Screen if an upgrade card is present when the AED Status Screen mode is launched.

**Before you begin:** Be sure the unit is in Maintenance Mode.

**To enter:** Navigate to **Upgrade AED:**

**AED Main Menu → AED maintenance → Upgrade AED**

**What it does:** If an upgrade data card is present, the unit will start performing the upgrade process. Follow any prompts and instructions that the upgrade application provides.



Do not turn off the unit or remove the battery pack or data card until the operation is complete.

**Note:** If the DDC card is not inserted, the unit will speak and display “Data Card Missing” (refer to the “*Installing the Defibtech Data Card (DDC Card)*” section in Chapter 3 of this manual). Press any softkey button to acknowledge the message and then install a Defibtech Data Card (DDC card).

**To exit:** When the unit finishes performing the AED upgrade, follow the displayed and spoken instructions.

#### ➔ Transfer Data To Card

**Transfer data to card** will initiate a data transfer from the DDU-2200 AED to a Defibtech Data Card (DDC card) inserted in the device. Internal event data and device history is written to the DDC card.

**Before you begin:** Be sure the unit is in Maintenance Mode. Be sure a DDC Card is installed in the device (refer to the “*Installing the Defibtech Data Card (DDC Card)*” section in Chapter 3 of this manual).

**To enter:** Navigate to **Transfer data to card:**

**AED Main Menu → AED maintenance → Transfer data to card**

**What it does:** The unit begins transferring rescue data to the card:

The unit says: “Transferring data to data card”

The unit displays: Transferring Data

The unit will finish transferring data to the data card and will speak and display: “data transfer complete.”



Do not turn off the unit or remove the battery pack or data card until the operation is complete.

**Note:** If a data card is not inserted, the unit will speak and display “Data Card Missing” (refer to the “*Installing the Defibtech Data Card (DDC Card)*” section in Chapter 3 of this manual).

**To exit:** When the unit finishes transferring the data to the data card, it will automatically return to the AED Maintenance menu screen.

#### ➔ Format Data Card

**Format data card** is a maintenance tool used to repair corrupted cards. It is unnecessary to perform this step on cards purchased with your DDU-2200 AED.



This step will erase all data on the data card!

**Before you begin:** Be sure the unit is in Maintenance Mode. Be sure a Defibtech Data Card (DDC card) is currently installed in the device (refer to the “*Installing the Defibtech Data Card (DDC Card)*” section in Chapter 3 of this manual).

**To enter:** Navigate to **Format data card**:

**AED Main Menu → AED maintenance → Format data card**

**What it does:** The unit will format the DDC card that is inserted in the AED:

The unit says: "Formatting data card"

The unit displays: Formatting Data Card

When unit finishes formatting the DDC card, the unit will return to the menu.



Do not turn off the unit or remove the battery pack or data card until the operation is complete.

**Note:** If the data card is not inserted, the unit will speak and display "Data Card Missing" (refer to the "Installing the Defibtech Data Card (DDC Card)" section in Chapter 3 of this manual).

**To exit:** When the unit finishes formatting the data card, it will automatically return to the AED Maintenance menu screen.



Using non-Defibtech Data Cards (DDC cards) may damage the unit and will void the warranty.

#### ➔ Run Application From Card

**Run application from card** will initiate a card application on the Defibtech Data Card (DDC card). The most typical application is a software upgrade.

**Before you begin:** Be sure the unit is in Maintenance Mode. Be sure a DDC Card, with a card application, is installed in the device (refer to the "Installing the Defibtech Data Card (DDC Card)" section in Chapter 3 of this manual).

**To enter:** Navigate to **Run application from card**:

**AED Main Menu → AED maintenance → Run application from card**



Do not turn off the unit or remove the battery pack or data card until the operation is complete.

**Note:** If a data card is not inserted, the unit will speak and display "Data Card Missing" (refer to the "Installing The Defibtech Data Card (DDC Card)" section in Chapter 3 of this manual).

**To exit:** When the unit finishes performing the application, follow the displayed and spoken instructions.

### 6.7 AED Options Screen

To manually configure AED options such as time, date, volume, and audio recording, select the **AED options** from the AED Main Menu screen.

**Before you begin:** Be sure the unit is in Maintenance Mode.

**To enter:** Navigate to **AED options**:

**AED Main Menu → AED options**

**What it does:** The unit will display the AED options menu screen. This screen will allow the user to modify the following user-changeable parameters:

- **System time**
- **System date**
- **Volume level**
- **Audio recording**
- **Status broadcast**

**To exit:** Using the TOP or BOTTOM softkey buttons, scroll to and highlight the selection.

**Go to main menu.** Press the CENTER softkey button. The unit will exit the AED options menu screen and return to the AED Main Menu.

### ➔ System Time

The **System time** option allows the user to set the time of the internal AED clock.

**Before you begin:** Be sure the unit is in Maintenance Mode.

**To enter:** Navigate to **System time**:

**AED Main Menu → AED options → System time**

**What it does:** The **System time** option allows the user to set the time of internal AED clock (using the 24 hour clock). Once the **System time** option is selected, press the CENTER softkey button to enter set time mode:

The hours selection will be highlighted green:

- Press the TOP or BOTTOM softkey buttons to adjust the hours to the desired time.
- Press the CENTER softkey button to accept the hours setting.

The minutes selection will be highlighted green:

- Press the TOP or BOTTOM softkey buttons to adjust the minutes to the desired time.
- Press the CENTER softkey button to accept the minutes setting.

The seconds selection will be highlighted green:

- Press the TOP or BOTTOM softkey buttons to adjust the seconds to the desired time.
- Press the CENTER softkey button to accept the seconds setting.

The time is now set and the user may use TOP or BOTTOM softkey buttons to navigate to other choices in the menu.

**Note:** The factory default setting of the internal AED clock is Universal Time (GMT).

### ➔ System Date

The **System date** option allows the user to set the date of the internal AED clock.

**Before you begin:** Be sure the unit is in Maintenance Mode.

**To enter:** Navigate to **System date**:

**AED Main Menu → AED options → System date**

**What it does:** The **System date** option allows the user to set the date of internal AED clock. Once the **System date** option is selected, press the CENTER softkey button to enter set date mode:

The year selection will be highlighted green:

- Press the TOP or BOTTOM softkey buttons to adjust the year.
- Press the CENTER softkey button to accept the year setting.

The month selection will be highlighted green:

- Press the TOP or BOTTOM softkey buttons to adjust the month.
- Press the CENTER softkey button to accept the month setting.

The day selection will be highlighted green:

- Press the TOP or BOTTOM softkey buttons to adjust the day.
- Press the CENTER softkey button to accept the day setting.



The date is now set and the user may use TOP or BOTTOM softkey buttons to navigate to other choices in the menu.

**Note:** The factory default setting of the internal AED clock is Universal Time (GMT).

### ➔ Volume Level

The **Volume level** option allows the user to set AED audio output to **high, medium, or low** volume. Changing the volume level will not change the volume of the Active Status Indicator “beep.”

**Before you begin:** Be sure the unit is in Maintenance Mode.

**To enter:** Navigate to **Volume level:**

**AED Main Menu → AED options → Volume level**

**What it does:** The **Volume level** option allows the user to set AED audio to **high, medium or low** volume. Once the **Volume level** option is selected, use the TOP and BOTTOM softkey buttons to cycle through the different volume settings. When the desired volume selection is chosen, press the CENTER softkey button to set that volume level. The AED will now use that volume setting for all audio (except the volume of the Active Status Indicator “beep”). The user may use the TOP or BOTTOM softkey buttons to navigate to other choices in the menu.

**Note:** The factory default setting of the volume level is “**High**.”

### ➔ Audio Recording

The **Audio recording** option allows the user to enable or disable recording of event audio data to a Defibtech Data Card (DDC card).

**Before you begin:** Be sure the unit is in Maintenance Mode.

**To enter:** Navigate to **Audio recording:**

**AED Main Menu → AED options → Audio recording**

**What it does:** The **Audio recording** option allows the user to enable/disable recording of event audio data. Once the **Audio recording** option item is selected, use the TOP and BOTTOM softkey buttons to select either enabled or disabled settings. When the desired selection is chosen, press the CENTER softkey button to set the feature. The AED will now use that audio recording setting. The user may use the TOP or BOTTOM softkey buttons to navigate to other choices in the menu.

**Note:** The factory default setting of the audio recording is “**Disabled**.”

### ➔ Status Broadcast

The **Status broadcast** option allows the user to enable or disable wireless (RF) transmission of status data to an optional data receiver (contact your Defibtech distributor for details).

**Before you begin:** Be sure the unit is in Maintenance Mode.

**To enter:** Navigate to **Status broadcast:**

**AED Main Menu → AED options → Status broadcast**

**What it does:** The **Status broadcast** option allows the user to enable/disable wireless (RF) transmission of status data to an optional data receiver. Once the **Status broadcast** option item is selected, use the TOP and BOTTOM softkey buttons to select either enabled or disabled settings. When the desired selection is chosen, press the CENTER softkey button to set the feature. The AED will now use that status broadcast setting. The user may use the TOP or BOTTOM softkey buttons to navigate to other choices in the menu.

**Note:** The factory default setting of the status broadcast is “**Disabled**.”

## 6.8 Rescue Options Screen

To manually configure rescue options such as Rescue protocol and CPR breathing, select the **Rescue options** from the AED Main Menu screen.

**Before you begin:** Be sure the unit is in Maintenance Mode.

**To enter:** Navigate to **Rescue options**:

**AED Main Menu → Rescue options**

**What it does:** The unit will display the Rescue options menu screen. This screen will allow the user to modify certain user changeable parameters:

- **CPR breathing**
- **Rescue protocol**
  - **Settings**

**To exit:** Using the TOP or BOTTOM softkey buttons, scroll to and highlight the selection **Go to main menu**. Press the CENTER softkey button. The unit will exit the Rescue options menu screen and return to the AED Main Menu.

### ➔ CPR Breathing

The **CPR breathing** option allows the user to set the AED's default CPR setting to include or not include CPR breathing coaching prompts during CPR.

**Before you begin:** Be sure the unit is in Maintenance Mode.

**To enter:** Navigate to **CPR breathing**:

**AED Main Menu → Rescue options → CPR breathing**

**What it does:** **CPR breathing** allows the user to set the AED's default CPR setting to include or not include CPR breathing coaching prompts.

Use the TOP and BOTTOM softkey buttons to select the desired mode. When the desired selection is chosen, press the CENTER softkey button to set the feature. The AED will now use that coaching setting.

**Note:** The factory default setting of CPR breathing is set to "**Disabled**."

### ➔ Rescue Protocol

The AED supports two rescue protocols at one time. The **Rescue protocol** option allows the user to select a rescue protocol. Rescue protocol options include the AHA/ERC protocol or "Custom"

**Before you begin:** Be sure the unit is in Maintenance Mode.

**To enter:** Navigate to **Rescue protocol**:

**AED Main Menu → Rescue options → Rescue protocol**

**What it does:** The **Rescue protocol** option allows the user to select between up to two rescue protocols that have been enabled in the unit. The factory default setting of the rescue protocol is "**AHA**."

To change the protocol, press the CENTER softkey button to highlight the protocol. The user will be prompted to enter a password to proceed. The password may be obtained from your medical director or from Defibtech (for Defibtech contact information, please refer to the "*Contacts*" section in Chapter 12). Once the password has been entered, the user may select between the two protocols.

To enter the password use the TOP softkey button to scroll through numbers. Once the correct number appears, use the CENTER softkey button to move to the next space. Once all of the numbers have been entered press the CENTER softkey button. The user will now be able to choose a different rescue protocol.

➔ **Settings**

**Before you begin:** Be sure the unit is in Maintenance Mode.

**To enter:** Navigate to **Settings**:

**AED Main Menu → Rescue options → Settings**

**What it does:** The **Settings** option allows the user to change the currently enabled protocol by entering a special protocol code. This code is a special code that encodes all the important information regarding the protocol (i.e. number of compressions and breaths per cycle, compression pacing, number of CPR cycles, and number of shocks between CPR). This code is custom generated by Defibtech. If the code is not entered correctly the protocol will not be changed. Based on the protocol code entered, the currently selected protocol will be changed to that described by the special protocol code. You can obtain this code from your medical director or from Defibtech (for Defibtech contact information, please refer to the “*Contacts*” section in Chapter 12). Once the code has been entered, the settings will have been changed.

To enter the code use the TOP softkey button to scroll through numbers/letters. Once the correct number/letter appears, use the CENTER softkey button to move to the next space. Once all of the numbers/letters have been entered, press the CENTER softkey button. The settings will have been changed based on the code entered.

## 6.9 Help Topics Screen

The **Help topics** option on the AED Main Menu provides a list of available help topics.

**Before you begin:** Be sure the unit is in Maintenance Mode.

**To enter:** Navigate to **Help topics**:

**AED Main Menu → Help topics**

**What it does:** The Help topics provides a list of available help topics.

The Help topics are the following:

- **Preparing the patient**
- **Analyzing and shocking**
- **Performing CPR**
- **Replacing the battery**
- **Replacing the pads**
- **Checking the AED status**
- **Replacing the data card**

Use the TOP and BOTTOM softkey buttons to scroll through the different Help topics items. When the desired help selection is highlighted (with a box), press the CENTER softkey button to get more information.

**To exit:** Using the TOP or BOTTOM softkey buttons, scroll and highlight the selection **Go to main menu**. Press the CENTER softkey button. The unit will exit the Help Topics menu screen and return to the AED Main Menu.

## 7 DDU-2200 AED Accessories

This chapter describes the component parts and accessories that can be used with the Defibtech DDU-2200 AED. For contact information on obtaining replacement component parts and accessories, refer to Chapter 12 in this manual. For more information about accessories, please visit [www.defibtech.com](http://www.defibtech.com) or contact Defibtech or your distributor.

### 7.1 Defibrillation Pads

The DDU-2200 AED must be used with Defibtech self-adhesive defibrillation pads for adults or with child/infant pads for infants and children. These defibrillation pads serve two functions:

- Allow the unit to read the patient's electrocardiogram (ECG) rhythm.
- Deliver defibrillation energy to the patient if needed.

The Defibtech self-adhesive defibrillation pad assembly comes in a "leads-out" sealed package that allows the device to be stored with the pads connected to the AED. When the DDU-2200 AED is used, the operator needs only to turn the device on, remove the pad package, tear open the package, remove the pads from the blue liner, apply the pads to the patient, and administer care. The AED has a storage area in the back of the unit that allows for storage of a single sealed adult pads package.

### 7.2 Battery Packs

The DDU-2200 AED uses a lithium battery pack to provide the AED with a long shelf and standby life. The battery pack is inserted into the battery pack opening on the back of the AED and latches into place. Standard battery packs (discernible by their yellow plastic casing) are not rechargeable.

### 7.3 Data Cards

The DDU-2200 AED is designed to optionally use Defibtech Data Cards (DDC cards). The AED will operate with or without a DDC card, but if a DDC card is installed, additional event storage capacity is available.

The DDU-2200 AED accepts DDC cards capable of recording an assortment of data for a given period of time. The DDU-2200 allows the user to enable or disable recording of audio data (refer to the "AED Options Screen" section in Chapter 6 of this manual).

The DDC card is inserted into a slot behind the data card/USB port access door found on the side of the AED (refer to the "Installing the Defibtech Data Card (DDC Card)" section in Chapter 3 of this manual). A new event file is created on the DDC card each time the AED is turned on, and the following information is recorded:

- The time the AED was turned on.
- Other data such as: ECG data, time data, audio data (audio enabled card only), event milestones such as: motion detection, shock advice, shock delivery information.

Multiple events may be recorded on a single DDC card. If the DDC card becomes full, the AED will stop recording to the card, however, the most relevant event documentation for the current session will still be recorded internally.

Internally recorded event information can be downloaded for external review by inserting a blank DDC card into the unit and following the Data Download procedure (refer to the "Downloading the Internal Data Log" section in Chapter 8 of this manual).



Using non-Defibtech Data Cards (DDC cards) may damage the unit and will void the warranty.

### 7.4 USB Cable

An optional USB cable may be used with the DDU-2200 AED for connecting the AED to a personal computer running Defibtech maintenance software. The AED has a mini-USB connector located on the right side of the unit behind the data card/USB port access door.



Do not have a USB cable attached to the unit during a rescue.

# 8 Event Viewing

This chapter includes information about DefibView, Defibtech Data Cards (DDC cards), and downloading internal data logs.

## 8.1 DefibView

DefibView is a Windows-based software application that reads data stored on a DDC card or downloaded through the USB port and displays the data on a personal computer. DefibView serves the following primary functions:

- Enables emergency care personnel to review a cardiac episode from the time the AED was turned on and connected to the patient until the unit is turned off.
- Provides maintenance personnel with additional parameter information to assist in troubleshooting a device suspected of malfunctioning.

DefibView is a stand-alone software application. DefibView cannot be used while the AED is in operation, and its function is solely to support post-event review.



Not intended for clinical use. Information presented by DefibView should not be used for making clinical decisions.

## 8.2 Defibtech Data Cards (DDC Cards)

If a DDC card is installed in the unit, every time the DDU-2200 AED is turned on, the following information is recorded on a new file on the card:

- The time the AED was turned on.
- Other data such as: ECG data, time data, audio data (audio enabled cards only), event milestones such as: motion detection, shock advice, shock delivery information.

This information can be reviewed using the DefibView application.



Using non-Defibtech Data Cards (DDC cards) may damage the unit and will void the warranty.

## 8.3 Downloading the Internal Data Log

Regardless of whether a DDC card is installed in the unit, select information is recorded internally within the DDU-2200 AED. The information recorded is limited to:

- The time the AED was turned on.
- Other data such as event milestones (motion detection, shock advice, shock delivery information, etc.).
- Important ECG information.

**Note:** Audio data is not logged internally.

### Downloading the Internal Data Log Using the DDC Card

To download the internally logged information, perform the following procedure:

- Insert a DDC card into the unit.
- Turn the unit on in Status Display Mode by pressing the center softkey button.
- Press the tool icon to enter AED Maintenance Mode.
- From the AED Maintenance screen, select the **Transfer data to card** option.
- Allow the unit to write the contents of the internal log to the DDC card.

The DDU-2200 AED will write the contents of the internal log onto the DDC card. This information can then be reviewed using DefibView software.

### Downloading the Internal Data Log Using the USB Port

To download the internal data log using the USB port in the unit, connect a USB cable between the unit and a PC. Launch the DefibView software and follow USB download directions.



Do not operate the DDU-2200 AED in Rescue Mode while a USB cable is plugged into the unit.

# 9 Technical Specifications

## 9.1 Defibtech DDU-2200 AED

### General

| Category              | Specification   |
|-----------------------|---|
| Size                  | 7.3 x 9.5 x 2.3 inches (18.5 x 24 x 5.8 cm)   |
| Weight                | Less than 3 lbs (1.4 kg) (with battery)   |
| Power                 | Battery Pack (not rechargeable)   |
| Design standards      | Meets applicable requirements of <ul style="list-style-type: none"> <li>• IEC 60601-1</li> <li>• UL 60601-1</li> <li>• CAN/CSA C22.2 No.601.1-M90</li> <li>• IEC 60601-1-2</li> <li>• IEC 60601-2-4</li> <li>• AAMI DF80</li> </ul> |
| Device Classification | Internally powered with defibrillator-proof BF-type patient applied parts (per EN 60601-1)  |
| Patient safety        | All patient connections are electrically isolated   |
| Rescue protocol       | AHA/ERC (default); supports protocol updates by the user (password protected)   |


### Defibrillator

| Category  | Specification  |
|---|--|
| Waveform  | Impedance Compensated Biphasic Truncated Exponential   |
| Energy  | Adult: 150 Joules (nominal [ $\pm 15\%$ ] delivered into a 50 ohm load)<br>Child/Infant: 50 Joules (nominal [ $\pm 15\%$ ] delivered into a 50 ohm load)   |
| Charge control  | Automatic by Patient Analysis System   |
| Charge time   | 4 seconds or less (from shock advised)*<br>Charge time may increase at the end of battery life and for temperatures below 10°C.  |
| Charge time from the initiation of rhythm analysis to readiness for discharge | Meets or exceeds AAMI DF80 and IEC 60601-2-4 requirements  |
| Charge time measured from initially switching power on to charge ready        | Meets or exceeds AAMI DF80 and IEC 60601-2-4 requirements  |
| Charge complete indication  | <ul style="list-style-type: none"> <li>• SHOCK required indicator flashing</li> <li>• Unit announces imminent shock delivery (“Shocking in 3, 2, 1...”) and displays “Delivering Shock” text prompt</li> </ul>   |
| Shock delivery  | Fully automatic  |
| DISARM  | <ul style="list-style-type: none"> <li>• If Patient Analysis System decides rhythm is no longer shockable</li> <li>• If defibrillation pads are removed from patient or unplugged from unit</li> <li>• If operator presses and holds the ON/OFF button for approximately two seconds, device will disarm and turn off</li> </ul> |

\*Typical, new battery, at 25°C

### Waveform Specifications

The DDU-2200 AED delivers a 150 J (Adult) 50 J (Pediatric) Biphasic Truncated Exponential waveform to patients with impedances ranging from 25 to 180 ohms.



The waveform is adjusted to compensate for measured patient impedance. Nominal phase times and energy delivered are shown in the tables below.

#### Phase Times (DDP-2001 Adult Defibrillation Pads)

| Patient Impedance | Phase A Duration | Phase B Duration |
|-------------------|------------------|------------------|
| 25 Ω              | 2.8 ms           | 2.8 ms           |
| 50 Ω              | 4.1 ms           | 4.1 ms           |
| 75 Ω              | 7.2 ms           | 4.8 ms           |
| 100 Ω             | 9.0 ms           | 6.0 ms           |
| 125 Ω             | 12.0 ms          | 8.0 ms           |
| 150 Ω             | 12.0 ms          | 8.0 ms           |
| 175 Ω             | 12.0 ms          | 8.0 ms           |

#### Energy (DDP-2001 Adult Defibrillation Pads) (Nominal)

| Nominal Energy | Load Impedance |       |       |       |       |       |       |
|----------------|----------------|-------|-------|-------|-------|-------|-------|
|                | 25 Ω           | 50 Ω  | 75 Ω  | 100 Ω | 125 Ω | 150 Ω | 175 Ω |
| 150 J          | 153 J          | 151 J | 152 J | 151 J | 153 J | 146 J | 142 J |

#### Phase Time and Energy (DDP-2002 Pediatric Defibrillation Pads)

| Patient Impedance | Phase A Duration | Phase B Duration | Energy Delivered |
|-------------------|------------------|------------------|------------------|
| 25 Ω              | 4.1 ms           | 4.1 ms           | 35 J             |
| 50 Ω              | 5.8 ms           | 3.8 ms           | 47 J             |
| 75 Ω              | 5.8 ms           | 3.8 ms           | 51 J             |
| 100 Ω             | 7.2 ms           | 4.8 ms           | 53 J             |
| 125 Ω             | 7.2 ms           | 4.8 ms           | 52 J             |
| 150 Ω             | 9 ms             | 6 ms             | 53 J             |
| 175 Ω             | 9 ms             | 6 ms             | 51 J             |

**Environmental**

| <b>Category</b>  |  | <b>Specification</b>   |
|--|--|--|
| Operating/<br>Maintenance                                    | Temperature  | 0 – 50°C (32 – 122°F)  |
|  | Humidity   | 5% – 95% (non-condensing)  |
|  | One Hour Operating Temperature Limit (extreme cold)* | -20°C (-4°F)   |
|  | Air Pressure   | 700 to 1060 hPa (21 to 31 inHg)  |
| Standby/Storage/<br>Transport                                | Temperature  | 0 – 50°C (32 – 122°F)  |
|  | Humidity   | 5% – 95% (non-condensing)  |
|  | Air Pressure   | 500 to 1060 hPa (15 to 31 inHg)  |
| Altitude   |  | -150 to 4500 meters (-500 to 15,000 feet) per MIL-STD-810F 500.4 Procedure II  |
| Shock/Drop Abuse Tolerance                                   |  | MIL-STD-810F 516.5 Procedure IV 48 in, (1.2 meters), any edge, corner, or surface, in standby mode   |
| Vibration  |  | MIL-STD-810F 514.5 Category 20 (Ground)<br>RTCA/DO-160D, Section 8.8.2, Cat R, Zone 2, Curve G (Helicopter)<br>RTCA/DO-160D, Section 8, Cat H, Zone 2, Curves B & R (Jet Aircraft) |
| Sealing/Water Resistance                                     |  | IEC 60529 class IP55; Dust Protected, Protected against water jets (Battery pack installed)  |
| ESD and EMI (radiated and immunity)                          |  | Refer to Chapter 10 for details  |
| Radio Frequency Emissions Applicable Directive and Standards |  | R&TTE Directive 1999/5/EC<br>ETSI EN 300 220-2 V2.1.2 (2007-06)<br>ERC RECOMMENDATION 70-03<br>ETSI EN 301 489-3 V1.4.1 (2002-08)  |
| Aviation   |  | Meets RTCA/DO-160G, Section 21, RF Radiated Emissions, Category M  |

\*From room temperature to temperature extreme, one hour duration.

**Patient Analysis System**



The DDU-2200 AED assesses proper pad/patient contact by measuring the impedance between the pads. To measure this impedance, 8 and 16 kHz sine waves at 74  $\mu$ A peak-to-peak maximum current are applied to the patient. The DDU-2200 Patient Analysis System ensures that the patient impedance is within the proper range and analyzes the patient's ECG rhythm to determine whether a shock is required. On detection of a non-shockable rhythm, the user is prompted to perform CPR. For shockable rhythms, the AED automatically charges in preparation for shock delivery.

The patient analysis system identifies and removes artifacts from the patient's ECG signal. Artifacts may arise from a variety of sources, including: noise, patient motion, respiration, muscular contractions, and pacemakers. Artifact that is caused by the patient or electrical noise may interfere with accurate rhythm analysis. When this artifact is present, the AED will prompt the user to "Stop Motion" or "Stop Interference" until the ECG signal is free of noise and then proceed to analysis.



### Shockable Rhythm Criteria

When placed on a patient meeting the indications for use criteria, the DDU-2200 AED is designed to recommend a defibrillation shock when it detects proper pad impedance and one of the following:

|  |
|--|
| <p><b>Ventricular Fibrillation:</b> Peak-to-peak amplitude at least 200 <math>\mu</math>Volts.</p>  <p>Some very low amplitude or low frequency VF rhythms may not be interpreted as shockable.</p> <p><b>WARNING</b></p>   |
| <p><b>Ventricular Tachycardia</b> (Including ventricular flutter and polymorphic VT):<br/>Cardiac rhythm rate of at least 180 bpm and peak-to-peak amplitude at least 200 <math>\mu</math>Volts.</p>  <p>Some very low amplitude or low frequency VT rhythms may not be interpreted as shockable.</p> <p><b>WARNING</b></p> |

The DDU-2200 AED is designed to recommend *no* shock for all other rhythms, including Normal Sinus Rhythms, fine Ventricular Fibrillation (<200  $\mu$ Volts), and some slow Ventricular Tachycardia and Asystole.

### Patient Analysis System Performance (typical)

| Rhythm Class  | ECG Test Sample <sup>1</sup> Size | Algorithm Performance <sup>1</sup> |   | Specifications  |
|---|-----------------------------------|------------------------------------|---|---|
|   |                                   | Performance <sup>2</sup>           | 90% Lower Confidence Limit <sup>2</sup> |   |
| Shockable Rhythm – Ventricular Fibrillation                         | 227                               | >97%                               | >95%                                    | Meets or exceeds IEC-60601-2-4 requirements; meets the AAMI DF80 requirement and the AHA recommendation <sup>2</sup> of Sensitivity > 90%                     |
| Shockable Rhythm – Ventricular Tachycardia                          | 101                               | >98%                               | >95%                                    | Meets or exceeds IEC-60601-2-4 requirements; meets the AAMI DF80 requirement and the AHA recommendation <sup>2</sup> of Sensitivity > 75%                     |
| Non-Shockable Rhythm – Normal Sinus Rhythm                          | 213                               | 100%                               | 100%                                    | Meets or exceeds IEC-60601-2-4 requirements; meets the AAMI DF80 requirement of Specificity >95% and the AHA recommendation <sup>2</sup> of Specificity > 99% |
| Non-Shockable Rhythm – Asystole                                     | 113                               | 100%                               | 100%                                    | Meets or exceeds IEC-60601-2-4 requirements; meets the AAMI DF80 requirement and the AHA recommendation <sup>2</sup> of Specificity > 95%                     |
| Non-Shockable Rhythm – All other non-shockable rhythms <sup>3</sup> | 248                               | >99%                               | >98%                                    | Meets or exceeds IEC-60601-2-4 requirements; meets the AAMI DF80 requirement and the AHA recommendation <sup>2</sup> of Specificity > 95%                     |
| Intermediate Rhythm – Fine Ventricular Fibrillation                 | 31                                | >90%                               | N/A                                     | Report only <sup>2</sup>  |
| Intermediate Rhythm – Other Ventricular Tachycardia Sinusoidal      | 17                                | >40%                               | N/A                                     | Report only <sup>2</sup>  |
| Intermediate Rhythm – Other Ventricular Tachycardia Horizontal      | 9                                 | >65%                               | N/A                                     | Report only <sup>2</sup>  |

1. From Defibtech ECG Rhythm Databases.

2. Automatic External Defibrillators for Public Access Defibrillation: Recommendations for Specifying and Reporting Arrhythmia Analysis Algorithm Performance, Incorporating New Waveforms, and Enhancing Safety. American Heart Association (AHA) Task Force on Automatic External Defibrillation, Subcommittee on AED Safety and Efficacy. Circulation, 1997;95:1677-1682.

3. Other Non-Shockable Rhythms include A-Fib (AF), A-Flutter (AFL), Heart Block (HB), Premature Ventricular Contractions (PVC), Sinus Bradycardia (SB), Supra-Ventricular Tachycardia (SVT), and idioventricular rhythms.

## 9.2 Battery Packs

Use only Defibtech battery packs in the DDU-2200 AED.

### DBP-2003 and DBP-2013 Battery Packs

| Category                         | Specification  |
|----------------------------------|--|
| Model number                     | DBP-2003<br>DBP-2013 (Aviation; TSO C-142a)  |
| Main battery type                | 12VDC, 2800 mAh, Lithium/Manganese Dioxide.<br>Disposable, recyclable, non-rechargeable. |
| Capacity                         | 125 shocks or 8 hours of continuous operation.*  |
| Charge time                      | 4 seconds or less (from shock advised)*  |
| Standby-life (installed in unit) | 4 years*   |

\*Typical, new battery, at 25°C

## 9.3 Self-Adhesive Defibrillation Pads

Defibtech self-adhesive defibrillation pads have the following characteristics:

| Model number            | DDP-2001  | DDP-2002  |
|-------------------------|---|---|
| Type                    | Adult   | Child/Infant < 8 years<br>< 55 lbs. (25 kg)                   |
| Intended use            | Disposable  | Disposable  |
| Adhesion                | Self-adhesive   | Self-adhesive   |
| Active gel surface area | 12 inches <sup>2</sup> (77 cm <sup>2</sup> ) each (nominal) | 7.75 inches <sup>2</sup> (50 cm <sup>2</sup> ) each (nominal) |
| Cable/connector type    | Integrated  | Integrated  |
| Cable length            | 4 feet (122 cm) (typical)                                   | 4 feet (122 cm) (typical)                                     |
| Expiration date         | 2.5 years from date of manufacture                          | 2.5 years from date of manufacture                            |

## 9.4 Event Documentation

### Internal Event Record

Select ECG segments and rescue event parameters are recorded and can be downloaded to a removable data card.

### Removable Storage (optional)

Up to 30 hours of ECG and event data storage (no audio option) or up to 3 hours of audio (audio option). ECG and event storage on a removable data card. Actual length of storage is dependent on card capacity. Data card must already be installed at the time of event.

## 9.5 Defibtech Event Viewer

DefibView is a PC-based application program that allows review of ECG data and other patient and device performance parameters after an emergency event.

DefibView runs on various Windows platforms including Windows XP and newer versions. Minimum system requirements for adequate performance are as follows:

- Pentium 4 processor
- 512 MB system memory
- 1 GB free space on hard disk
- USB 1.0 connectivity

## 9.6 Recycling Information

---

At the end of useful life, recycle the defibrillator and its accessories.

### Recycling Assistance

For recycling assistance contact your local Defibtech distributor.  
Recycle in accordance with local and national regulations.

### Preparation For Recycling

Items should be clean and contaminant-free prior to being recycled.  
When recycling used disposable electrodes, follow local clinical procedures.

### Packaging For Recycling

Packaging should be recycled in accordance with local and national requirements.

## 9.7 Notice to European Union Customers

---



The crossed-out wheeled bin symbol on this device indicates that this equipment has been put on the market after 13 August 2005, and is included in the scope of the directive 2002/96/EEC on Waste Electrical and Electronic Equipment (WEEE) and of the national decree(s), which transpose provisions of such directive.

At the end of its lifetime, this device can only be disposed of in compliance with the provisions of the above-mentioned European directive (and as amended) as well as with the corresponding national regulations. Severe penalties are possible for unauthorized disposal.

Electrical and Electronic Equipment (EEE) may contain polluting components and hazardous substances, the accumulation of which could pose serious risk for the environment and human health. It is for this reason that local administrations provide regulations, which encourage reuse and recycling, and prohibit the disposal of WEEE as unsorted municipal waste and require the collection of such WEEE separately (at specifically authorized treatment facilities). Manufacturers and authorized distributors are required to supply information about a safe treatment and disposition of the specific device.

You may also return this equipment to your distributor when purchasing a new one. As for reuse and recycling, notwithstanding the limits imposed by the nature and the use of this device, the manufacturer will do its best to develop recovery processes. Please contact the local distributor for information.

# 10 Electromagnetic Conformity

## 10.1 Guidance and Manufacturer's Declaration

The essential performance of the DDU-2200 AED is successful delivery of defibrillation therapy and accurate differentiation between shockable and nonshockable rhythms.

The DDU-2200 AED is intended for use within the electromagnetic environment specified below. The customer or the user of the DDU-2200 AED should assure that it is used within the stated environmental specifications.


### ELECTROMAGNETIC EMISSIONS

| <b>Emissions test</b>                                | <b>Compliance</b>                        | <b>Electromagnetic environment – guidance</b>   |
|--|--|---|
| RF emissions<br>CISPR 11<br>CISPR 22<br>FCC Part 15  | Group 1<br>Class B<br>Class B<br>Class B | The DDU-2200 AED uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment. |
| Harmonic emissions IEC 61000-3-2                     | Not applicable                           | Battery operated equipment  |
| Voltage fluctuations/flicker emissions IEC 61000-3-3 | Not applicable                           | Battery operated equipment  |

### ELECTROMAGNETIC IMMUNITY

| <b>Immunity test</b>  | <b>IEC 60601 test level</b>                                       | <b>Compliance level</b>     | <b>Electromagnetic environment – guidance</b>   |
|---|---|-----------------------------|---|
| Electrostatic discharge (ESD)<br>IEC 61000-4-2  | ±8 kV contact<br>±15 kV air                                       | ±8 kV contact<br>±15 kV air | There are no special requirements with respect to electrostatic discharge.  |
| Electrical fast transient/burst<br>IEC 61000-4-4  | ±2 kV for power line supply lines<br>±1 kV for input/output lines | Not applicable              | Battery operated equipment  |
| Surge<br>IEC 61000-4-5  | ±1 kV line(s) to line(s)<br>±2 kV line(s) to earth                | Not applicable              | Battery operated equipment  |
| Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11 | Not applicable  | Not applicable              | Battery operated equipment  |
| Power frequency (50/60 Hz) magnetic field IEC 61000-4-8   | 3 A/m   | 3 A/m                       | Power frequency magnetic fields should not be greater than levels characteristic of a typical location in a commercial or hospital environment. |

Electromagnetic Immunity (continued)

| Immunity test   | IEC 60601 test level   | Compliance level | Electromagnetic environment – guidance   |
|---|--|------------------|--|
| Radiated RF<br>IEC 61000-4-3  | 20 V/m<br>80 MHz to<br>2.5 GHz<br><br>80% 5Hz AM<br>Modulation | 20 V/m           | Portable and mobile RF communications equipment should be used no closer to any part of the DDU-2200 AED, including cables, than necessary. The recommended separation distance calculated from the equation applicable to the frequency of the transmitter is shown in the following table. |
|   |  |                  |  Interference may occur in the vicinity of equipment marked with this symbol.  |
| <p><b>Note 1:</b> At 80 MHz and 800 MHz, the higher frequency range applies.</p> <p><b>Note 2:</b> These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.</p> <p>The ISM (industrial, scientific and medical) bands between 150 kHz and 80 MHz are 6,765 MHz to 6,795 MHz; 13,553 MHz to 13,567 MHz; 26,957 MHz to 27,283 MHz; and 40,66 MHz to 40,70 MHz.</p> <p>Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the DDU-2200 AED is used exceeds the applicable RF compliance level above, the DDU-2200 AED should be observed to verify normal operation. If abnormal performance is observed additional measures may be necessary, such as reorienting or relocating the DDU-2200 AED.</p> |  |                  |  |

**Separation Distances**

The DDU-2200 AED is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of a DDU-2200 AED can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the DDU-2200 AED as recommended below, according to the maximum output of the communications equipment.

| Recommended separation distances between portable and mobile RF communications equipment and DDU-2200 AEDs |   |                    |
|--|---|--------------------|
| Rated maximum output power of transmitter  | Separation distance according to frequency of transmitter |                    |
|  | 80 MHz to 800 MHz   | 800 MHz to 2.5 GHz |
|  | $d = 1.2\sqrt{P}$   | $d = 2.3\sqrt{P}$  |
| 0.01 W   | 0.12 m  | 0.23 m             |
| 0.1 W  | 0.38 m  | 0.73 m             |
| 1 W  | 1.20 m  | 2.30 m             |
| 10 W   | 3.79 m  | 7.27 m             |
| 100 W  | 12.00 m   | 23.00 m            |

*Separation Distances (continued)*

For transmitters rated at a maximum output power not listed above, the recommended separation distance (d) in meters (m) can be determined using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

**Note 1:** As 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.

**Note 2:** The ISM (industrial, scientific and medical) bands between 150 kHz and 80 MHz are 6,765 MHz to 6,795 MHz; 13,553 MHz to 13,567 MHz; 26,957 MHz to 27,283 MHz; and 40,66 MHz to 40,70 MHz.

**Note 3:** An additional factor of 10/3 is used in calculating the recommended separation distance for transmitters in the ISM frequency bands between 150 kHz and 80 MHz and in the frequency range 80 MHz to 2.5 GHz to decrease the likelihood that mobile/portable communications equipment could cause interference if it is inadvertently brought into patient areas.

**Note 4:** These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

**Regulatory Compliance**

Changes or modifications of this product, not expressly approved by Defibtech, may void the user's authority to operate the equipment.

This device complies with part 15 of the FCC Rules and Industry Canada Radio Standard RSS-210. Operation is subject to the following two conditions:

- (1) This device may not cause harmful interference, and
- (2) This device must accept any interference received, including interference that may cause undesired operation.

















This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.


















**CE Marking and European Union Compliance – Radio Transmitter**

Defibtech, L.L.C. declares that the DDU-2200 AED radio transmitter is in compliance with the essential requirements and other relevant provisions of Directive 1999/5/EC. Applicable standards are listed in the "Environmental" section in Chapter 9 of this manual.

# 11 Glossary of Symbols

| Symbol  | Meaning  |
|---|--|
|    | High voltage present.  |
|    | SHOCK Required Indicator – Flashes to indicate that a shock is about to be delivered.  |
|    | ON/OFF Button<br><ul style="list-style-type: none"> <li>• Turns the device ON when it is OFF.</li> <li>• Turns the device OFF when it is ON.</li> </ul>  |
|    | Caution, consult accompanying documents.   |
|    | Do not expose to high heat or open flame. Do not incinerate.   |
|    | Recyclable.  |
|    | Consult operating instructions.  |
|    | Refer to instruction manual / booklet.   |
|  | Do not damage or crush.  |
|  | Follow proper disposal procedures.   |
|  | Meets the requirements of the European Medical Device Directive.   |
|  | Meets the requirements of the Radio Equipment and Telecommunications Directive, 1999/5/EC.   |
|  | Classified by TUV Rheinland of NA with respect to electric shock, fire, and mechanical hazard only in accordance with UL 60601-1, CAN/CSA C22.2 No.601.1-M90, IEC 60601-1, and IEC 60601-2-4. Conforms to UL Standard UL 60601-1. Certified to CAN/CSA Standard C22.2 No. 601.1-M90. |
|  | Authorized European Representative:<br>EMERGO EUROPE<br>Prinsessegracht 20<br>2514 AP The Hague<br>The Netherlands   |
|  | Operational temperature limitation.  |
|  | Use by (yyyy-mm-dd).   |

Glossary Of Symbols (continued)

| Symbol  | Meaning   |
|---|---|
|    | Defibrillation proof - Can withstand the effects of an externally applied defibrillation shock. Internally powered with defibrillator-proof BF-type patient applied parts (per EN 60601-1). |
|    | Manufacturer.   |
|    | Date of manufacture.  |
|    | Manufacturer and date of manufacture.   |
|    | Do not reuse.   |
|    | For USA users only.   |
| Rx ONLY   | Federal Law (USA) restricts this device to sale by or on the order of a physician.  |
|    | Catalogue number.   |
|   | Keep dry.   |
|  | Handle with care.   |
|  | Transportation and storage requirements.<br>See environmental requirements on packaging.  |
|  | Does not contain latex.   |
|  | Lot number.   |
| IP55  | Dust protected; Protected against water jets.   |
|  | Serial number.  |
|  | Lithium manganese dioxide battery.  |
|  | Lithium-ion battery.  |
|  | Product is not sterile.   |
|  | Defibrillation proof - Can withstand the effects of an externally applied defibrillation shock. Internally powered with defibrillator-proof CF-type patient applied parts (per EN 60601-1). |



# 12 Contacts

## Manufacturer



Defibtech, L.L.C.  
741 Boston Post Road, Suite 201  
Guilford, CT 06437 USA

Tel.: 1-(866) 333-4241 (Toll-free within North America)  
1-(203) 453-4507  
Fax: 1-(203) 453-6657

Email:  
sales@defibtech.com (Sales)  
reporting@defibtech.com (Medical Device Reporting)  
service@defibtech.com (Service and Repair)

This product and its accessories are manufactured and sold under license to at least one or more of the following United States patents: 5,591,213; 5,593,427; 5,601,612; 5,607,454; 5,611,815; 5,617,853; 5,620,470; 5,662,690; 5,735,879; 5,749,904; 5,749,905; 5,776,166; 5,800,460; 5,803,927; 5,836,978; 5,836,993; 5,879,374; 6,016,059; 6,047,212; 6,075,369; 6,438,415; 6,441,582.

For additional patent information, please see:  
[www.defibtech.com/patents](http://www.defibtech.com/patents)

