

# Lifeline AUTO



## Lifeline AUTO Fully Automatic Defibrillator

Defibtech Lifeline AEDs offer industry-leading innovation, simplicity, and elegance. The Lifeline AUTO is a fully automated defibrillator designed to analyze heart rhythm and automatically deliver a shock - without intervention - if needed.

Technologically advanced enough to include all mission critical features necessary to provide the most advanced treatment for Sudden Cardiac Arrest. Yet so simple and unintimidating to use that virtually anyone can quickly and effectively help save a life.

The Lifeline AUTO was developed by experienced multidisciplinary engineering teams and incorporates state-of-the-art digital signal processing techniques and advanced ECG analysis algorithms. The device exceeds the American Heart Association performance

recommendations - giving the user confidence the correct therapy is being delivered.

In the event of a cardiac arrest, defibrillation using an AED within three minutes can increase survival rates from less than 8 percent to more than 70 percent. The Lifeline AUTO defibrillator, with its one-button operation, uses advanced biphasic technology and automatically adjusts the shock delivery to the person's individual needs.

Defibtech AED deployments include workplaces, government buildings, airports and aircraft, rail stations and trains, educational institutions, malls, factories, emergency vehicles, healthcare facilities, resorts, arenas, and marine vessels. More than a quarter-million units have been shipped worldwide.

# Defibtech Lifeline AUTO Fully Automatic Defibrillator

## TECHNICAL SPECIFICATIONS†

### DEFIBRILLATOR

#### TYPE

Fully automatic external defibrillator

#### MODEL

DDU-120E

#### WAVEFORM

Biphasic Truncated Exponential (Impedance compensated)

#### ENERGY

Adult: 150 Joules  
Child / Infant: 50 Joules  
(Nominal into 50 Ohm load)

#### CHARGE TIME\*

4 seconds or less  
(from shock advised)

#### VOICE PROMPTS

Extensive voice prompts guide user through operation of the unit

#### CPR PACING

Metronome

#### CONTROLS

Lighted On/Off button

#### INDICATORS

- “check pads”
- “do not touch patient”
- “analyzing”
- armed-for-shock
- AED Status LED

#### RESCUE PROTOCOL

AHA/ERC (default); supports protocol updates

\*Typical, with new battery at 25°C

### PATIENT ANALYSIS SYSTEM

#### PATIENT ANALYSIS

Automatically evaluates patient impedance for proper pad contact. Monitors signal quality and analyzes patient ECG for shockable/non-shockable rhythms.

#### SENSITIVITY/SPECIFICITY

Meets or exceeds IEC-60601-2-4 requirements; meets AAMI DF80 requirements and AHA recommendations

### BATTERY PACK

MODEL DBP-2800	MODEL DBP-1400
<b>POWER</b> 15VDC, 2800 mAh	<b>POWER</b> 15VDC, 1400 mAh
<b>CAPACITY</b> 300 shocks or 16 hours continuous operation*	<b>CAPACITY</b> 125 shocks or 8 hours continuous operation*
<b>STANDBY LIFE</b> 7 years (installed in AED with 9V ASI battery)*	<b>STANDBY LIFE</b> 5 years (installed in AED with 9V ASI battery)*

#### TYPE

Lithium/Manganese Dioxide  
Disposable, recyclable, non-rechargeable

#### LOW BATTERY INDICATORS

Visible  
Audible

\*Typical, with new battery at 25°C

### SELF TESTS

#### AUTOMATIC

Automatic daily, weekly, monthly and quarterly circuitry tests

#### BATTERY INSERTION

System integrity test on battery insertion

#### PAD PRESENCE

Pads preconnected tested daily

#### USER-INITIATED

Unit and battery pack system test initiated by the user

#### STATUS INDICATION

Visual and audible indication of unit status

### DEFIBRILLATION / MONITORING PADS

#### MODEL

Adult: DDP-100  
Child / Infant: DDP-200P

#### TYPE

Pre-connected, single-use, non-polarized, disposable, self-adhesive electrodes with cable and connector

#### SURFACE AREA\*\*

Adult: 16 inches<sup>2</sup> (103 cm<sup>2</sup>)  
Child / Infant: 7.75 inches<sup>2</sup> (50 cm<sup>2</sup>)

\*\*Nominal, each pad

### EVENT DOCUMENTATION

#### INTERNAL EVENT RECORD

Critical ECG segments and rescue event parameters are recorded and can be downloaded to a removable data card

#### PC-BASED EVENT REVIEW

ECG with event tag display, and audio playback when available

#### REMOVABLE STORAGE

(optional) Up to 12 hours of ECG and event data storage (no audio option) or up to 1 hour and 40 minutes of audio (audio option). ECG and event storage on a removable data card. Actual length of storage is dependent on card capacity.

### ENVIRONMENTAL

#### TEMPERATURE

Operating: 0 to 50°C (32 to 122°F)  
One Hour Operating Temperature Limit (extreme cold): -20°C (-4°F)\*\*\*

Standby: 0 to 50°C (32 to 122°F)

#### RELATIVE HUMIDITY

Operating / Standby: 5%-95% (non-condensing)

#### ALTITUDE

-500 to 15,000 ft (-150 to 4500 m) per MIL-STD-810F 500.4 Procedure II

#### VIBRATION

Ground (MIL-STD-810F 514.5 Category 20)

Helicopter (RTCA/DO-160D, Section 8.8.2, Cat R, Zone 2, Curve G)

Jet Aircraft (RTCA/DO-160D Section 8, Cat H, Zone 2, Curves B & R)

#### SHOCK / DROP ABUSE TOLERANCE

MIL-STD-810F 516.5 Procedure IV (1 meter, any edge, corner, or surface, in standby mode)

#### SEALING / WATER RESISTANCE

IEC 60529 class IP54; Dust Protected, Splash Proof (battery pack installed)

#### ESD

EN 61000-4-2 (15kV or direct contact up to 8kV)

#### EMC (Emission)

EN 55011 Class B Group 1 and FCC Part 15

#### EMC (Immunity)

EN 61000-4-3 (20V/m)

\*\*\*From room temperature to temperature extreme, one hour duration

### SELF TESTS

#### AUTOMATIC

Automatic daily, weekly, monthly and quarterly circuitry tests

#### BATTERY INSERTION

System integrity test on battery insertion

#### PAD PRESENCE

Pads preconnected tested daily

#### USER-INITIATED

Unit and battery pack system test initiated by the user

#### STATUS INDICATION

Visual and audible indication of unit status

### PHYSICAL

#### SIZE

8.5 x 11.8 x 2.7 inches  
(22 x 30 x 7 cm)

#### WEIGHT (Approximate)

With DBP-1400: 4.2 lbs (1.9 kg)  
With DBP-2800: 4.4 lbs (2.0 kg)



†Specifications subject to change without notice



Defibtech, LLC • Guilford, CT 06437 USA • 1-203-453-4507 • 1-866-DEFIB-4U (1-866-333-4248)  
www.defibtech.com

DAC-E705-EN-BB

ELECTRONIC DISTRIBUTION

Issued: 2018-10-22